

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071118\
 Data File : VW003886.D
 Acq On : 12 Jul 2018 01:18
 Operator : SY/AP
 Sample : J3914-09RE
 Misc : 4.80G/5ML/MSVOA W/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SEMHRE

Quant Time: Jul 12 03:43:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 07 05:07:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	164345	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	268345	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	187003	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	51878	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	114690	53.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.72%	
35) Dibromofluoromethane	7.88	113	100508	53.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.04%	
50) Toluene-d8	10.33	98	308745	42.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.96%	
62) 4-Bromofluorobenzene	12.62	95	63917	24.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	49.18%	
Target Compounds						
16) Acetone	4.14	43	11927	31.987	ug/l	99
20) Methylene Chloride	4.90	84	23461	9.346	ug/l	96
64) Tetrachloroethene	10.87	164	4579	3.310	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW071118\
Data File : VW003886.D
Acq On : 12 Jul 2018 01:18
Operator : SY/AP
Sample : J3914-09RE
Misc : 4.80G/5ML/MSVOA W/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SEMHRE

Quant Time: Jul 12 03:43:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 07 05:07:25 2018
Response via : Initial Calibration

