

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071118\
 Data File : VW003889.D
 Acq On : 12 Jul 2018 02:35
 Operator : SY/AP
 Sample : J3918-01RE
 Misc : 4.88G/5ML/MSVOA W/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VNJ-204-255-ETSRE

Quant Time: Jul 12 04:09:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 07 05:07:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	12068	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	19268	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	19735	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	8251	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	9737	61.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.40%	
35) Dibromofluoromethane	7.89	113	7957	59.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.00%	
50) Toluene-d8	10.33	98	21732	42.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.26%	
62) 4-Bromofluorobenzene	12.62	95	8100	43.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.80%	
Target Compounds						
16) Acetone	4.12	43	6049	220.923	ug/l	98
20) Methylene Chloride	4.92	84	5526	40.587	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW071118\
Data File : VW003889.D
Acq On : 12 Jul 2018 02:35
Operator : SY/AP
Sample : J3918-01RE
Misc : 4.88G/5ML/MSVOA W/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VNJ-204-255-ETSRE

Quant Time: Jul 12 04:09:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 07 05:07:25 2018
Response via : Initial Calibration

