

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071119\
 Data File : VW011198.D
 Acq On : 10 Jul 2019 18:42
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 11 08:20:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070819S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 11 08:14:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.373	0.338	9.4	86	0.00
3 P	Chloromethane	0.481	0.460	4.4	92	0.00
4 C	Vinyl Chloride	0.629	0.599	4.8#	87	0.00
5 T	Bromomethane	0.316	0.317	-0.3	91	0.00
6 T	Chloroethane	0.339	0.338	0.3	89	0.00
7 T	Trichlorofluoromethane	0.223	0.198	11.2	80	0.00
8 T	Diethyl Ether	0.309	0.310	-0.3	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.535	0.528	1.3	89	0.00
10 T	Methyl Iodide	0.736	0.738	-0.3	89	0.00
11 T	Tert butyl alcohol	0.047	0.044	6.4	89	0.00
12 CM	1,1-Dichloroethene	0.555	0.554	0.2#	90	0.00
13 T	Acrolein	0.037	0.045	-21.6#	111	0.00
14 T	Allyl chloride	1.013	0.975	3.8	84	0.00
15 T	Acrylonitrile	0.157	0.159	-1.3	92	0.00
16 T	Acetone	0.151	0.154	-2.0	93	0.00
17 T	Carbon Disulfide	1.684	1.631	3.1	86	0.00
18 T	Methyl Acetate	0.473	0.381	19.5	91	0.00
19 T	Methyl tert-butyl Ether	0.843	0.846	-0.4	90	0.00
20 T	Methylene Chloride	0.596	0.606	-1.7	93	0.00
21 T	trans-1,2-Dichloroethene	0.577	0.580	-0.5	90	0.00
22 T	Diisopropyl ether	2.179	2.244	-3.0	92	0.00
23 T	Vinyl Acetate	1.348	1.468	-8.9	92	0.00
24 P	1,1-Dichloroethane	1.144	1.162	-1.6	91	0.00
25 T	2-Butanone	0.237	0.240	-1.3	93	0.00
26 T	2,2-Dichloropropane	0.595	0.558	6.2	84	0.00
27 T	cis-1,2-Dichloroethene	0.623	0.635	-1.9	91	0.00
28 T	Bromochloromethane	0.520	0.477	8.3	85	0.00
29 T	Tetrahydrofuran	0.156	0.158	-1.3	94	0.00
30 C	Chloroform	1.059	1.064	-0.5#	92	0.00
31 T	Cyclohexane	1.217	1.147	5.8	88	0.00
32 T	1,1,1-Trichloroethane	0.806	0.794	1.5	87	0.00
33 S	1,2-Dichloroethane-d4	0.628	0.618	1.6	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.298	0.297	0.3	93	0.00
36 T	1,1-Dichloropropene	0.532	0.530	0.4	88	0.00
37 T	Ethyl Acetate	0.320	0.329	-2.8	94	0.00
38 T	Carbon Tetrachloride	0.451	0.453	-0.4	89	0.00
39 T	Methylcyclohexane	0.645	0.654	-1.4	88	0.00
40 TM	Benzene	1.471	1.493	-1.5	90	0.00
41 T	Methacrylonitrile	0.183	0.200	-9.3	104	0.00
42 TM	1,2-Dichloroethane	0.477	0.491	-2.9	93	0.00
43 T	Isopropyl Acetate	0.604	0.625	-3.5	93	0.00
44 TM	Trichloroethene	0.354	0.354	0.0	90	0.00
45 C	1,2-Dichloropropane	0.398	0.414	-4.0#	92	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071119\
 Data File : VW011198.D
 Acq On : 10 Jul 2019 18:42
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 11 08:20:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070819S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 11 08:14:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.187	0.192	-2.7	92	0.00
47 T	Bromodichloromethane	0.479	0.497	-3.8	92	0.00
48 T	Methyl methacrylate	0.291	0.302	-3.8	92	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	95	0.00
50 S	Toluene-d8	1.205	1.203	0.2	91	0.00
51 T	4-Methyl-2-Pentanone	0.316	0.330	-4.4	95	0.00
52 CM	Toluene	0.893	0.918	-2.8#	90	0.00
53 T	t-1,3-Dichloropropene	0.510	0.519	-1.8	90	0.00
54 T	cis-1,3-Dichloropropene	0.600	0.609	-1.5	89	0.00
55 T	1,1,2-Trichloroethane	0.266	0.277	-4.1	95	0.00
56 T	Ethyl methacrylate	0.404	0.427	-5.7	94	0.00
57 T	1,3-Dichloropropane	0.502	0.524	-4.4	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.187	0.203	-8.6	101	0.00
59 T	2-Hexanone	0.219	0.230	-5.0	95	0.00
60 T	Dibromochloromethane	0.291	0.305	-4.8	93	0.00
61 T	1,2-Dibromoethane	0.252	0.261	-3.6	94	0.00
62 S	4-Bromofluorobenzene	0.442	0.438	0.9	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.337	0.328	2.7	90	0.00
65 PM	Chlorobenzene	1.017	1.023	-0.6	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.358	0.363	-1.4	93	0.00
67 C	Ethyl Benzene	1.946	1.961	-0.8#	91	0.00
68 T	m/p-Xylenes	0.704	0.715	-1.6	91	0.00
69 T	o-Xylene	0.664	0.676	-1.8	92	0.00
70 T	Styrene	1.145	1.185	-3.5	93	0.00
71 P	Bromoform	0.190	0.187	1.6	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.738	3.738	0.0	91	0.00
74 T	N-amyl acetate	1.302	1.338	-2.8	96	0.00
75 P	1,1,2,2-Tetrachloroethane	0.766	0.764	0.3	95	0.00
76 T	1,2,3-Trichloropropane	0.578	0.536	7.3	95	0.00
77 T	Bromobenzene	0.839	0.836	0.4	93	0.00
78 T	n-propylbenzene	4.604	4.623	-0.4	91	0.00
79 T	2-Chlorotoluene	2.650	2.658	-0.3	92	0.00
80 T	1,3,5-Trimethylbenzene	3.175	3.203	-0.9	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.255	0.239	6.3	87	0.00
82 T	4-Chlorotoluene	2.760	2.747	0.5	92	0.00
83 T	tert-Butylbenzene	2.648	2.675	-1.0	92	0.00
84 T	1,2,4-Trimethylbenzene	3.188	3.228	-1.3	93	0.00
85 T	sec-Butylbenzene	3.821	3.837	-0.4	91	0.00
86 T	p-Isopropyltoluene	3.427	3.477	-1.5	91	0.00
87 T	1,3-Dichlorobenzene	1.614	1.656	-2.6	96	0.00
88 T	1,4-Dichlorobenzene	1.625	1.637	-0.7	93	0.00
89 T	n-Butylbenzene	3.419	3.539	-3.5	92	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071119\
Data File : VW011198.D
Acq On : 10 Jul 2019 18:42
Operator : SY/VA
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Jul 11 08:20:15 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070819S.M
Quant Title : SW846 8260
QLast Update : Thu Jul 11 08:14:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.626	0.620	1.0	90	0.00
91 T	1,2-Dichlorobenzene	1.464	1.479	-1.0	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.136	0.132	2.9	93	0.00
93 T	1,2,4-Trichlorobenzene	1.008	1.076	-6.7	96	0.00
94 T	Hexachlorobutadiene	0.619	0.639	-3.2	92	0.00
95 T	Naphthalene	1.932	2.102	-8.8	95	0.00
96 T	1,2,3-Trichlorobenzene	0.890	0.965	-8.4	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6