

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071218\
 Data File : VW003910.D
 Acq On : 12 Jul 2018 20:59
 Operator : SY/AP
 Sample : J3947-05
 Misc : 3.89G/5ML/MSVOA W/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 PAR-3

Quant Time: Jul 13 06:58:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 07 05:07:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	138350	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	221710	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	178974	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	54467	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	95692	52.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.78%	
35) Dibromofluoromethane	7.88	113	86096	55.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.96%	
50) Toluene-d8	10.33	98	296479	49.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.90%	
62) 4-Bromofluorobenzene	12.62	95	74442	34.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	69.32%	
Target Compounds						
16) Acetone	4.15	43	45969	146.446	ug/l	97
17) Carbon Disulfide	4.37	76	12858	3.018	ug/l	98
20) Methylene Chloride	4.92	84	9341	1.887	ug/l	91
25) 2-Butanone	7.19	43	13067	26.578	ug/l	92
80) 1,3,5-Trimethylbenzene	12.95	105	3475	1.093	ug/l	95
84) 1,2,4-Trimethylbenzene	13.26	105	10443	3.162	ug/l	99
95) Naphthalene	15.38	128	271022	134.157	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW071218\
Data File : VW003910.D
Acq On : 12 Jul 2018 20:59
Operator : SY/AP
Sample : J3947-05
Misc : 3.89G/5ML/MSVOA W/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
PAR-3

Quant Time: Jul 13 06:58:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070718S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 07 05:07:25 2018
Response via : Initial Calibration

