

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071221\
 Data File : VW019424.D
 Acq On : 12 Jul 2021 11:35
 Operator : SY/VA
 Sample : M2970-01
 Misc : 5.23G/10ML/MSVOA_W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGE76

Quant Time: Jul 13 02:21:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jul 13 02:19:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	450890	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	403810	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	162960	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	88136	13.796	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	55.200%
7) Chloroethane-d5	2.898	69	74181	16.283	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.120%
11) 1,1-Dichloroethene-d2	4.025	63	145449	11.800	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	47.200%
21) 2-Butanone-d5	7.092	46	54330	29.301	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	58.600%
24) Chloroform-d	7.653	84	210798	16.920	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	67.680%
26) 1,2-Dichloroethane-d4	8.311	65	113151	16.550	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	66.200%#
32) Benzene-d6	8.281	84	426007	17.723	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	70.880%
36) 1,2-Dichloropropane-d6	9.274	67	131019	17.488	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	69.960%#
41) Toluene-d8	10.323	98	367301	16.948	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	67.800%
43) trans-1,3-Dichloroprop...	10.579	79	46821	14.593	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	58.360%
47) 2-Hexanone-d5	10.927	63	41099	29.569	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	59.140%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	93257	15.405	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	61.600%
66) 1,2-Dichlorobenzene-d4	13.853	152	93306	16.228	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	64.920%#
Target Compounds						
					Qvalue	
13) Acetone	4.147	43	56666	41.475	ug/L	88
14) Carbon disulfide	4.391	76	6656	0.353	ug/L #	88
16) Methylene chloride	4.928	84	31233	3.475	ug/L	97
22) 2-Butanone	7.189	43	19431	9.191	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071221\
Data File : VW019424.D
Acq On : 12 Jul 2021 11:35
Operator : SY/VA
Sample : M2970-01
Misc : 5.23G/10ML/MSVOA_W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGE76

Quant Time: Jul 13 02:21:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070921SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jul 13 02:19:34 2021
Response via : Initial Calibration

