

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071221\
 Data File : VW019442.D
 Acq On : 12 Jul 2021 20:37
 Operator : SY/VA
 Sample : M2970-16
 Misc : 4.81G/10ML/MSVOA_W/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGE17

Quant Time: Jul 13 04:38:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jul 13 04:35:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	493403	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	434390	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	186476	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	110111	15.751	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	63.000%
7) Chloroethane-d5	2.897	69	96682	19.394	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.560%
11) 1,1-Dichloroethene-d2	4.025	63	186575	13.832	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	55.320%
21) 2-Butanone-d5	7.080	46	58800	28.979	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	57.960%
24) Chloroform-d	7.653	84	264460	19.398	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	77.600%
26) 1,2-Dichloroethane-d4	8.311	65	139852	18.693	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	74.760%
32) Benzene-d6	8.281	84	537551	20.789	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	83.160%
36) 1,2-Dichloropropane-d6	9.274	67	168102	20.858	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	83.440%
41) Toluene-d8	10.323	98	473127	20.294	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	81.160%
43) trans-1,3-Dichloroprop...	10.579	79	55745	16.151	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	64.600%
47) 2-Hexanone-d5	10.926	63	47239	31.594	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	63.180%
56) 1,1,2,2-Tetrachloroeth...	12.694	84	109378	16.796	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	67.200%
66) 1,2-Dichlorobenzene-d4	13.853	152	123821	18.820	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	75.280%
Target Compounds						Qvalue
13) Acetone	4.135	43	12173	8.142	ug/L	87
16) Methylene chloride	4.921	84	28893	2.937	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071221\
Data File : VW019442.D
Acq On : 12 Jul 2021 20:37
Operator : SY/VA
Sample : M2970-16
Misc : 4.81G/10ML/MSVOA_W/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGE17

Quant Time: Jul 13 04:38:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070921SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jul 13 04:35:18 2021
Response via : Initial Calibration

