

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071221\
 Data File : VW019448.D
 Acq On : 12 Jul 2021 23:01
 Operator : SY/VA
 Sample : M2973-08
 Misc : 4.07G/10ML/MSVOA_W/SOIL
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGEC3

Quant Time: Jul 13 04:39:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jul 13 04:35:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	463920	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	384450	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	134389	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	120215	18.289	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	73.160%		
7) Chloroethane-d5	2.891	69	105317	22.468	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	89.880%		
11) 1,1-Dichloroethene-d2	4.025	63	200460	15.806	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	63.240%		
21) 2-Butanone-d5	7.080	46	75000	39.312	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	78.620%		
24) Chloroform-d	7.653	84	296160	23.104	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	92.400%		
26) 1,2-Dichloroethane-d4	8.305	65	165411	23.514	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	94.040%		
32) Benzene-d6	8.275	84	590651	25.810	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	103.240%		
36) 1,2-Dichloropropane-d6	9.274	67	189669	26.591	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	106.360%		
41) Toluene-d8	10.323	98	505817	24.515	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	98.080%		
43) trans-1,3-Dichloroprop...	10.579	79	55667	18.224	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	72.880%		
47) 2-Hexanone-d5	10.926	63	64405	48.670	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	97.340%		
56) 1,1,2,2-Tetrachloroeth...	12.694	84	134730	23.376	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	93.520%		
66) 1,2-Dichlorobenzene-d4	13.853	152	108518	22.886	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	91.560%		
Target Compounds					Qvalue	
13) Acetone	4.129	43	10021	7.128	ug/L	89
16) Methylene chloride	4.928	84	34441	3.724	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071221\
Data File : VW019448.D
Acq On : 12 Jul 2021 23:01
Operator : SY/VA
Sample : M2973-08
Misc : 4.07G/10ML/MSVOA_W/SOIL
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGEC3

Quant Time: Jul 13 04:39:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070921SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jul 13 04:35:18 2021
Response via : Initial Calibration

