

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071619\
 Data File : VW011290.D
 Acq On : 16 Jul 2019 17:06
 Operator : SY/VA
 Sample : K3798-07RE
 Misc : 5.30G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A52H4RE

Quant Time: Jul 17 08:41:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM070219S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 17 07:20:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	266226	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	218845	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	73338	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	60837	14.96	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	59.84%
7) Chloroethane-d5	2.89	69	50723	17.78	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.12%
10) 1,1-Dichloroethene-d2	4.01	63	112417	13.06	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	52.24%
20) 2-Butanone-d5	7.07	46	107554	67.17	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	134.34%
24) Chloroform-d	7.65	84	158196	21.68	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.72%
26) 1,2-Dichloroethane-d4	8.31	65	110361	24.56	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	98.24%
29) Benzene-d6	8.27	84	310882	23.06	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.24%
33) 1,2-Dichloropropane-d6	9.27	67	111792	25.46	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.84%
37) Toluene-d8	10.32	98	251428	20.93	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	83.72%
38) trans-1,3-Dichloropropene-	10.58	79	42405	22.72	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	90.88%
39) 2-Hexanone-d5	10.92	63	71489	71.01	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	142.02%#
48) 1,1,2,2-Tetrachloroethane-	12.69	84	107494	29.60	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	118.40%
61) 1,2-Dichlorobenzene-d4	13.85	152	63137	22.43	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.72%

Target Compounds

					Ovalue
13) Acetone	4.12	43	87256	55.787	ug/L 99
15) Methyl Acetate	4.67	43	28587	11.232	ug/L 98
21) 2-Butanone	7.18	43	28486	14.108	ug/L 99
44) Toluene	10.38	91	619298	43.745	ug/L 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071619\
Data File : VW011290.D
Acq On : 16 Jul 2019 17:06
Operator : SY/VA
Sample : K3798-07RE
Misc : 5.30G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A52H4RE

Quant Time: Jul 17 08:41:43 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM070219S.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 17 07:20:32 2019
Response via : Initial Calibration

