

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071718\  
 Data File : VW004008.D  
 Acq On : 18 Jul 2018 03:27  
 Operator : VA/AP  
 Sample : J3821-01RE  
 Misc : 5.75G/5ML/MSVOA W/SOIL  
 ALS Vial : 39 Sample Multiplier: 1

Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 HR-05-071218-ARE

Quant Time: Jul 18 04:53:57 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_W\METHOD\82W071618S.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jul 16 15:06:12 2018  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 7.95  | 168  | 154103   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 8.85  | 114  | 267195   | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 11.63 | 117  | 59331    | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 13.57 | 152  | 64328    | 50.00 | ug/l  | 0.00     |

| System Monitoring Compounds | R.T.  | QIon | Response | Conc  | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|-------|---------|----------|
| 33) 1,2-Dichloroethane-d4   | 8.31  | 65   | 118782   | 68.96 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 137.92% |          |
| 35) Dibromofluoromethane    | 7.88  | 113  | 89228    | 52.05 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 104.10% |          |
| 50) Toluene-d8              | 10.33 | 98   | 43194    | 6.49  | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 12.98%  |          |
| 62) 4-Bromofluorobenzene    | 12.63 | 95   | 49200    | 20.51 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 41.02%  |          |


| Target Compounds               | R.T.  | QIon | Response | Conc      | Units  | Qvalue |
|--------------------------------|-------|------|----------|-----------|--------|--------|
| 16) Acetone                    | 4.12  | 43   | 24674    | 59.609    | ug/l   | 97     |
| 18) Methyl Acetate             | 4.68  | 43   | 3452     | 3.474     | ug/l   | 99     |
| 20) Methylene Chloride         | 4.92  | 84   | 9104     | Below Cal |        | 98     |
| 25) 2-Butanone                 | 7.18  | 43   | 4233     | 7.599     | ug/l # | 77     |
| 30) Chloroform                 | 7.67  | 83   | 3410     | 1.125     | ug/l   | 88     |
| 31) Cyclohexane                | 7.96  | 56   | 3868     | 1.354     | ug/l # | 23     |
| 36) 1,1-Dichloropropene        | 7.95  | 75   | 12946    | 4.709     | ug/l # | 49     |
| 37) Ethyl Acetate              | 7.18  | 43   | 4233     | 3.192     | ug/l # | 70     |
| 38) Carbon Tetrachloride       | 7.95  | 117  | 15689    | 6.108     | ug/l # | 18     |
| 56) Ethyl methacrylate         | 10.66 | 69   | 84       | 2.985     | ug/l # | 27     |
| 68) m/p-Xylenes                | 11.84 | 106  | 1199     | 1.415     | ug/l   | 85     |
| 69) o-Xylene                   | 12.18 | 106  | 1067     | 1.358     | ug/l   | 99     |
| 76) 1,2,3-Trichloropropane     | 12.62 | 75   | 26329    | 41.227    | ug/l # | 100    |
| 78) n-propylbenzene            | 12.81 | 91   | 6035     | 1.082     | ug/l   | 91     |
| 80) 1,3,5-Trimethylbenzene     | 12.95 | 105  | 14005    | 3.661     | ug/l   | 98     |
| 81) trans-1,4-Dichloro-2-buten | 12.62 | 75   | 26329    | 91.194    | ug/l # | 11     |
| 84) 1,2,4-Trimethylbenzene     | 13.26 | 105  | 45260    | 11.500    | ug/l   | 96     |
| 85) sec-Butylbenzene           | 13.39 | 105  | 6885     | 1.445     | ug/l # | 74     |
| 86) p-Isopropyltoluene         | 13.51 | 119  | 8597     | 2.057     | ug/l   | 92     |
| 89) n-Butylbenzene             | 13.83 | 91   | 20561    | 5.004     | ug/l # | 65     |
| 90) Hexachloroethane           | 13.89 | 117  | 1655     | 2.148     | ug/l   | 61     |
| 92) 1,2-Dibromo-3-Chloropropan | 14.51 | 75   | 243      | 1.555     | ug/l   | 77     |
| 95) Naphthalene                | 15.38 | 128  | 6891     | 2.746     | ug/l # | 94     |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW071718\  
Data File : VW004008.D  
Acq On : 18 Jul 2018 03:27  
Operator : VA/AP  
Sample : J3821-01RE  
Misc : 5.75G/5ML/MSVOA W/SOIL  
ALS Vial : 39 Sample Multiplier: 1

Instrument :  
MSVOA\_W  
ClientSampleId :  
HR-05-071218-ARE

Quant Time: Jul 18 04:53:57 2018  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_W\METHOD\82W071618S.M  
Quant Title : SW846 8260  
QLast Update : Mon Jul 16 15:06:12 2018  
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.95 min Scan# 1044

Delta R.T. 0.00 min

Lab File: VW004008.D

Acq: 18 Jul 2018 03:27


Instrument :  
MSVOA\_W  
ClientSampleId :  
HR-05-071218-ARE

Tot Ion:168 Resp: 154103

Ion Ratio Lower Upper

168 100

99 57.8 43.8 65.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VW

7.95

80000

60000



40000

20000

0

7.90 8.00

Time-->



#16

Acetone

Concen: 59.609 ug/l

RT: 4.12 min Scan# 416

Delta R.T. 0.00 min

Lab File: VW004008.D


Acq: 18 Jul 2018 03:27

Tgt Ion: 43 Resp: 24674

Ion Ratio Lower Upper

43 100

58 28.9 24.5 36.7



Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 58.00 (57.70 to 58.70): VW

4.12

10000

8000

6000


4000


2000

0

4.00 4.10 4.20

Time-->









#18  
Methyl Acetate  
Concen: 3.474 ug/l  
RT: 4.68 min Scan# 507  
Delta R.T. 0.01 min  
Lab File: VW004008.D  
Acq: 18 Jul 2018 03:27




Instrument :  
MSVOA\_W  
ClientSampleId :  
HR-05-071218-ARE


Tot Ion: 43 Resp: 3452  
Ion Ratio Lower Upper  
43 100  
74 22.5 17.7 26.5



#20  
Methylene Chloride  
Concen: Below Cal  
RT: 4.92 min Scan# 546  
Delta R.T. 0.01 min  
Lab File: VW004008.D  
Acq: 18 Jul 2018 03:27

Tgt Ion: 84 Resp: 9104  
Ion Ratio Lower Upper  
84 100  
49 129.9 102.9 154.3  
51 43.9 30.9 46.3  
86 63.3 49.9 74.9









#25  
2-Butanone  
Concen: 7.599 ug/l  
RT: 7.18 min Scan# 918  
Delta R.T. 0.01 min  
Lab File: VW004008.D  
Acq: 18 Jul 2018 03:27




Instrument :  
MSVOA\_W  
ClientSampleId :  
HR-05-071218-ARE


Tot Ion: 43 Resp: 4233  
Ion Ratio Lower Upper  
43 100  
72 13.3 19.8 29.6#



#30  
Chloroform  
Concen: 1.125 ug/l  
RT: 7.67 min Scan# 998  
Delta R.T. -0.01 min  
Lab File: VW004008.D  
Acq: 18 Jul 2018 03:27

Tgt Ion: 83 Resp: 3410  
Ion Ratio Lower Upper  
83 100  
85 55.7 52.2 78.4





#31

Cyclohexane

Concen: 1.354 ug/l

RT: 7.96 min Scan# 1045

Delta R.T. 0.01 min

Lab File: VW004008.D

Acq: 18 Jul 2018 03:27

Instrument :

MSVOA\_W

ClientSampleId :

HR-05-071218-ARE


Tot Ion: 56 Resp: 3868

Ion Ratio Lower Upper

56 100

69 136.8 27.8 41.8#




84 116.1 67.2 100.8#



Abundance Ion 56.00 (55.70 to 56.70): VW

Ion 69.00 (68.70 to 69.70): VW

Ion 84.00 (83.70 to 84.70): VW



#33

1,2-Dichloroethane-d4

Concen: 68.957 ug/l

RT: 8.31 min Scan# 1103

Delta R.T. 0.01 min

Lab File: VW004008.D


Acq: 18 Jul 2018 03:27

Tgt Ion: 65 Resp: 118782

Ion Ratio Lower Upper



65 100


67 52.0 0.0 105.0



Abundance Ion 64.90 (64.60 to 65.60): VW

Ion 66.90 (66.60 to 67.60): VW






#34  
 1,4-Difluorobenzene  
 Concen: 50.000 ug/l  
 RT: 8.85 min Scan# 1191  
 Delta R.T. 0.00 min  
 Lab File: VW004008.D  
 Acq: 18 Jul 2018 03:27

Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 HR-05-071218-ARE




| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 114     | 100   |       |       |
| 63      | 22.0  | 0.0   | 42.8  |
| 88      | 16.2  | 0.0   | 30.4  |



Abundance Ion 113.90 (113.60 to 114.60): V


Ion 63.00 (62.70 to 63.70): VW

Ion 87.90 (87.60 to 88.60): VW



#35  
 Dibromofluoromethane  
 Concen: 52.047 ug/l  
 RT: 7.88 min Scan# 1033  
 Delta R.T. 0.00 min  
 Lab File: VW004008.D  
 Acq: 18 Jul 2018 03:27



| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 113     | 100   |       |       |
| 111     | 102.5 | 81.3  | 121.9 |
| 192     | 19.6  | 15.8  | 23.6  |




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#36  
 1,1-Dichloropropene  
 Concen: 4.709 ug/l  
 RT: 7.95 min Scan# 1044  
 Delta R.T. -0.13 min  
 Lab File: VW004008.D  
 Acq: 18 Jul 2018 03:27




Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 HR-05-071218-ARE


| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 75      | 100   |       |       |
| 110     | 7.3   | 17.0  | 51.0# |
| 77      | 0.0   | 24.9  | 37.3# |



#37  
 Ethyl Acetate  
 Concen: 3.192 ug/l  
 RT: 7.18 min Scan# 918  
 Delta R.T. -0.07 min  
 Lab File: VW004008.D  
 Acq: 18 Jul 2018 03:27

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 43      | 100   |       |       |
| 61      | 0.0   | 10.7  | 16.1# |
| 70      | 0.0   | 7.7   | 11.5# |









#38  
Carbon Tetrachloride  
Concen: 6.108 ug/l  
RT: 7.95 min Scan# 1044  
Delta R.T. -0.12 min  
Lab File: VW004008.D  
Acq: 18 Jul 2018 03:27




Instrument :  
MSVOA\_W  
ClientSampleId :  
HR-05-071218-ARE


Tot Ion:117 Resp: 15689  
Ion Ratio Lower Upper  
117 100  
119 6.3 75.3 112.9#  
121 0.0 26.0 39.0#



#50  
Toluene-d8  
Concen: 6.493 ug/l  
RT: 10.33 min Scan# 1434  
Delta R.T. 0.00 min  
Lab File: VW004008.D  
Acq: 18 Jul 2018 03:27

Tgt Ion: 98 Resp: 43194  
Ion Ratio Lower Upper  
98 100  
100 64.4 51.2 76.8





#56

Ethyl methacrylate

Concen: 2.985 ug/l

RT: 10.66 min Scan# 1488

Delta R.T. 0.01 min

Lab File: VW004008.D

Acq: 18 Jul 2018 03:27

Instrument :

MSVOA\_W

ClientSampleId :

HR-05-071218-ARE





Tot Ion: 69 Resp: 84

Ion Ratio Lower Upper

69 100

41 0.0 56.5 84.7#

39 67.9 30.2 45.4#



#62

4-Bromofluorobenzene

Concen: 20.510 ug/l

RT: 12.63 min Scan# 1811

Delta R.T. 0.01 min

Lab File: VW004008.D

Acq: 18 Jul 2018 03:27




Tgt Ion: 95 Resp: 49200


Ion Ratio Lower Upper

95 100

174 81.3 0.0 160.2

176 78.0 0.0 155.8









#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.63 min Scan# 1648  
Delta R.T. 0.00 min  
Lab File: VW004008.D  
Acq: 18 Jul 2018 03:27




Instrument :  
MSVOA\_W  
ClientSampleId :  
HR-05-071218-ARE


Tot Ion:117 Resp: 59331  
Ion Ratio Lower Upper  
117 100  
82 53.8 46.5 69.7  
119 32.0 26.2 39.2



#68  
m/p-Xylenes  
Concen: 1.415 ug/l  
RT: 11.84 min Scan# 1682  
Delta R.T. 0.00 min  
Lab File: VW004008.D  
Acq: 18 Jul 2018 03:27

Tgt Ion:106 Resp: 1199  
Ion Ratio Lower Upper  
106 100  
91 182.3 164.2 246.2









#69  
o-Xylene  
Concen: 1.358 ug/l  
RT: 12.18 min Scan# 1737  
Delta R.T. 0.01 min  
Lab File: VW004008.D  
Acq: 18 Jul 2018 03:27

Instrument :  
MSVOA\_W  
ClientSampleId :  
HR-05-071218-ARE

Tot Ion:106 Resp: 1067  
Ion Ratio Lower Upper  
106 100  
91 213.8 107.5 322.6




Abundance Ion 106.00 (105.70 to 106.70): V  
Ion 91.00 (90.70 to 91.70): VV






#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.57 min Scan# 1965  
Delta R.T. 0.00 min  
Lab File: VW004008.D  
Acq: 18 Jul 2018 03:27

Tgt Ion:152 Resp: 64328  
Ion Ratio Lower Upper  
152 100  
115 60.7 29.2 87.6  
150 157.7 0.0 345.2



Abundance Ion 151.90 (151.60 to 152.60): V  
Ion 114.90 (114.60 to 115.60): V  
Ion 149.90 (149.60 to 150.60): V





#76

1,2,3-Trichloropropane

Concen: 41.227 ug/l

RT: 12.62 min Scan# 1810

Delta R.T. -0.15 min

Lab File: VW004008.D

Acq: 18 Jul 2018 03:27

Instrument :

MSVOA\_W

ClientSampleId :





HR-05-071218-ARE

Tot Ion: 75 Resp: 26329

Ion Ratio Lower Upper

75 100

77 1.7 0.0 0.0#



#78

n-propylbenzene

Concen: 1.082 ug/l

RT: 12.81 min Scan# 1841

Delta R.T. 0.00 min

Lab File: VW004008.D




Acq: 18 Jul 2018 03:27


Tgt Ion: 91 Resp: 6035

Ion Ratio Lower Upper

91 100

120 26.9 11.2 33.6





#80

1,3,5-Trimethylbenzene

Concen: 3.661 ug/l

RT: 12.95 min Scan# 1864

Delta R.T. 0.00 min

Lab File: VW004008.D

Acq: 18 Jul 2018 03:27

Instrument :

MSVOA\_W

ClientSampleId :


HR-05-071218-ARE

Tot Ion:105 Resp: 14005

Ion Ratio Lower Upper




105 100

120 49.9 24.4 73.2



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#81

trans-1,4-Dichloro-2-butene

Concen: 91.194 ug/l

RT: 12.62 min Scan# 1810

Delta R.T. 0.10 min

Lab File: VW004008.D

Acq: 18 Jul 2018 03:27


Tgt Ion: 75 Resp: 26329

Ion Ratio Lower Upper

75 100

53 0.3 77.6 116.4#



89 0.2 36.6 54.8#




Abundance Ion 74.90 (74.60 to 75.60): VW

Ion 53.00 (52.70 to 53.70): VW

Ion 88.90 (88.60 to 89.60): VW









#84  
 1,2,4-Trimethylbenzene  
 Concen: 11.500 ug/l  
 RT: 13.26 min Scan# 1914  
 Delta R.T. 0.00 min  
 Lab File: VW004008.D  
 Acq: 18 Jul 2018 03:27




Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 HR-05-071218-ARE


Tot Ion:105 Resp: 45260  
 Ion Ratio Lower Upper  
 105 100  
 120 46.5 22.1 66.1



#85  
 sec-Butylbenzene  
 Concen: 1.445 ug/l  
 RT: 13.39 min Scan# 1936  
 Delta R.T. 0.00 min  
 Lab File: VW004008.D  
 Acq: 18 Jul 2018 03:27

Tgt Ion:105 Resp: 6885  
 Ion Ratio Lower Upper  
 105 100  
 134 31.3 9.8 29.4#









#86  
 p-Isopropyltoluene  
 Concen: 2.057 ug/l  
 RT: 13.51 min Scan# 1955  
 Delta R.T. 0.00 min  
 Lab File: VW004008.D  
 Acq: 18 Jul 2018 03:27

Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 HR-05-071218-ARE

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 119     | 100   |       |       |
| 134     | 27.3  | 13.0  | 38.9  |
| 91      | 30.2  | 11.8  | 35.3  |




Abundance Ion 119.00 (118.70 to 119.70): V  
 Ion 134.00 (133.70 to 134.70): V  
 Ion 91.00 (90.70 to 91.70): V






#89  
 n-Butylbenzene  
 Concen: 5.004 ug/l  
 RT: 13.83 min Scan# 2008  
 Delta R.T. 0.00 min  
 Lab File: VW004008.D  
 Acq: 18 Jul 2018 03:27

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 91      | 100   |       |       |
| 92      | 32.7  | 26.1  | 78.3  |
| 134     | 0.0   | 12.4  | 37.4# |



Abundance Ion 91.00 (90.70 to 91.70): V  
 Ion 92.00 (91.70 to 92.70): V  
 Ion 134.00 (133.70 to 134.70): V





#90

Hexachloroethane

Concen: 2.148 ug/l

RT: 13.89 min Scan# 2018

Delta R.T. -0.21 min

Lab File: VW004008.D

Acq: 18 Jul 2018 03:27

Instrument :

MSVOA\_W

ClientSampleId :





HR-05-071218-ARE

Tot Ion:117 Resp: 1655

Ion Ratio Lower Upper

117 100

201 46.3 40.3 120.9



#92

1,2-Dibromo-3-Chloropropane

Concen: 1.555 ug/l

RT: 14.51 min Scan# 2120

Delta R.T. 0.02 min

Lab File: VW004008.D

Acq: 18 Jul 2018 03:27




Tgt Ion: 75 Resp: 243


Ion Ratio Lower Upper

75 100

155 72.8 39.4 118.1

157 66.7 51.8 155.4





#95

Naphthalene

Concen: 2.746 ug/l

RT: 15.38 min Scan# 2262

Delta R.T. 0.00 min

Lab File: VW004008.D

Acq: 18 Jul 2018 03:27

Instrument :

MSVOA\_W

ClientSampleId :

HR-05-071218-ARE


Tot Ion:128 Resp: 6891

Ion Ratio Lower Upper

128 100

127 10.2 10.7 16.1#

129 9.6 8.7 13.1



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

