

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071718\
 Data File : VW003995.D
 Acq On : 17 Jul 2018 19:39
 Operator : VA/AP
 Sample : J3997-07
 Misc : 6.85G/5ML/MSVOA W/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-F-D

Quant Time: Jul 18 05:44:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 16 15:06:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	157772	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	274490	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	263333	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	124948	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	111690	63.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	126.66%	
35) Dibromofluoromethane	7.88	113	94151	53.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.92%	
50) Toluene-d8	10.33	98	315277	46.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.26%	
62) 4-Bromofluorobenzene	12.63	95	107951	43.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.62%	
Target Compounds						
16) Acetone	4.12	43	23549	55.57	ug/l	Qvalue 97
30) Chloroform	7.68	83	8262	2.66	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW071718\
Data File : VW003995.D
Acq On : 17 Jul 2018 19:39
Operator : VA/AP
Sample : J3997-07
Misc : 6.85G/5ML/MSVOA W/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-F-D

Quant Time: Jul 18 05:44:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 16 15:06:12 2018
Response via : Initial Calibration

