

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071718\
 Data File : VW004007.D
 Acq On : 18 Jul 2018 03:01
 Operator : VA/AP
 Sample : J3820-03RE
 Misc : 6.89G/5ML/MSVOA W/SOIL
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BU-02-071218RE

Quant Time: Jul 18 07:15:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 16 15:06:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	154241	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	289565	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	234386	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	113333	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	139506	80.92	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	161.84%	
35) Dibromofluoromethane	7.89	113	118330	63.69	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	127.38%	
50) Toluene-d8	10.33	98	287580	39.89	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	79.78%	
62) 4-Bromofluorobenzene	12.62	95	144977	55.77	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	111.54%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.12	43	16828	40.62	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW071718\
Data File : VW004007.D
Acq On : 18 Jul 2018 03:01
Operator : VA/AP
Sample : J3820-03RE
Misc : 6.89G/5ML/MSVOA W/SOIL
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BU-02-071218RE

Quant Time: Jul 18 07:15:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 16 15:06:12 2018
Response via : Initial Calibration

