

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071718\
 Data File : VW004012.D
 Acq On : 18 Jul 2018 05:11
 Operator : VA/AP
 Sample : J3958-01RE
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 CS-STONERE

Quant Time: Jul 18 07:49:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 16 15:06:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	175250	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	325296	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	299651	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	144707	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	134243	68.53	ug/l	0.00
Spiked Amount			Recovery	=	137.06%	
35) Dibromofluoromethane	7.89	113	114491	54.85	ug/l	0.00
Spiked Amount			Recovery	=	109.70%	
50) Toluene-d8	10.33	98	404844	49.99	ug/l	0.00
Spiked Amount			Recovery	=	99.98%	
62) 4-Bromofluorobenzene	12.63	95	140025	47.95	ug/l	0.00
Spiked Amount			Recovery	=	95.90%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.13	43	4606	9.78	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071718\
Data File : VW004012.D
Acq On : 18 Jul 2018 05:11
Operator : VA/AP
Sample : J3958-01RE
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
CS-STONERE

Quant Time: Jul 18 07:49:34 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 16 15:06:12 2018
Response via : Initial Calibration

