

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071718\
 Data File : VW004021.D
 Acq On : 18 Jul 2018 09:03
 Operator : VA/AP
 Sample : J3997-07RE
 Misc : 6.51G/5ML/MSVOA W/SOIL
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-F-DRE

Quant Time: Jul 18 11:06:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 16 15:06:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	147050	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	272318	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	253009	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	112175	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	104181	63.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	126.76%	
35) Dibromofluoromethane	7.88	113	91947	52.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.24%	
50) Toluene-d8	10.33	98	308023	45.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.86%	
62) 4-Bromofluorobenzene	12.63	95	100396	41.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.14%	
Target Compounds						
16) Acetone	4.13	43	20380	51.60	ug/l	Qvalue 94
30) Chloroform	7.68	83	7620	2.63	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW071718\
Data File : VW004021.D
Acq On : 18 Jul 2018 09:03
Operator : VA/AP
Sample : J3997-07RE
Misc : 6.51G/5ML/MSVOA W/SOIL
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-F-DRE

Quant Time: Jul 18 11:06:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 16 15:06:12 2018
Response via : Initial Calibration

