

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071718\
 Data File : VW004023.D
 Acq On : 18 Jul 2018 09:55
 Operator : VA/AP
 Sample : J3997-15RE
 Misc : 4.74G/5ML/MSVOA W/SOIL
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-H-DRE

Quant Time: Jul 18 11:09:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 16 15:06:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	146560	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	278339	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	266151	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	127314	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	125396	76.54	ug/l	0.00
Spiked Amount			Recovery	=	153.08%	
35) Dibromofluoromethane	7.89	113	106691	59.74	ug/l	0.00
Spiked Amount			Recovery	=	119.48%	
50) Toluene-d8	10.33	98	364168	52.55	ug/l	0.00
Spiked Amount			Recovery	=	105.10%	
62) 4-Bromofluorobenzene	12.63	95	128775	51.53	ug/l	0.00
Spiked Amount			Recovery	=	103.06%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.12	43	21168	53.77	ug/l	99
20) Methylene Chloride	4.91	84	18307	3.17	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW071718\
Data File : VW004023.D
Acq On : 18 Jul 2018 09:55
Operator : VA/AP
Sample : J3997-15RE
Misc : 4.74G/5ML/MSVOA W/SOIL
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-H-DRE

Quant Time: Jul 18 11:09:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 16 15:06:12 2018
Response via : Initial Calibration

