

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW071718\
 Data File : VW004024.D
 Acq On : 18 Jul 2018 10:21
 Operator : VA/AP
 Sample : J4012-01
 Misc : 6.08G/5ML/MSVOA W/SOIL
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 FS-RT-4917

Quant Time: Jul 18 11:11:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 16 15:06:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	80134	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	128116	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	100180	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	33691	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	54307	60.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	121.26%	
35) Dibromofluoromethane	7.88	113	47561	57.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	115.72%	
50) Toluene-d8	10.33	98	102857	32.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	64.50%	
62) 4-Bromofluorobenzene	12.63	95	35167	30.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	61.14%	

Target Compounds						Qvalue
20) Methylene Chloride	4.91	84	11285	4.76	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW071718\
Data File : VW004024.D
Acq On : 18 Jul 2018 10:21
Operator : VA/AP
Sample : J4012-01
Misc : 6.08G/5ML/MSVOA W/SOIL
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
FS-RT-4917

Quant Time: Jul 18 11:11:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 16 15:06:12 2018
Response via : Initial Calibration

