


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW071819\
 Data File : VW011346.D
 Acq On : 18 Jul 2019 21:08
 Operator : SY/VA
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070819S.M
 Title : SW846 8260
 Last Update : Thu Jul 11 08:14:17 2019



AutoFind: Scans 1756, 1757, 1758; Background Corrected with Scan 1747

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.1	30568	PASS
75	95	30	60	54.0	71437	PASS
95	95	100	100	100.0	132376	PASS
96	95	5	9	6.8	8955	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	71.9	95232	PASS
175	174	5	9	8.3	7891	PASS
176	174	95	101	95.2	90661	PASS
177	176	5	9	7.4	6739	PASS