




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071822\
 Data File : VW023845.D
 Acq On : 18 Jul 2022 16:31
 Operator : SY/VA
 Sample : N3701-11
 Misc : 6.67g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 F5B84

Manual IntegrationsAPPROVED

Quant Time: Jul 19 01:09:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jul 19 01:05:58 2022
 Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 07/19/2022
 Supervised By :Mahesh Dadoda 07/20/2022



TIC: VW023845.D\data.ms

(63) 1,2,4-Trimethylbenzene

13.182min (-0.067) 0.01 ug/L

response 57

Ion	Exp%	Act%
105.00	100.00	100.00
120.00	44.40	49.12
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071822\
 Data File : VW023845.D
 Acq On : 18 Jul 2022 16:31
 Operator : SY/VA
 Sample : N3701-11
 Misc : 6.67g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 F5B84

Manual IntegrationsAPPROVED

Quant Time: Jul 19 01:09:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jul 19 01:05:58 2022
 Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 07/19/2022
 Supervised By :Mahesh Dadoda 07/20/2022



TIC: VW023845.D\data.ms

(63) 1,2,4-Trimethylbenzene

13.249min (-0.000) 0.83 ug/L m

response 7833

Ion	Exp%	Act%
105.00	100.00	100.00
120.00	44.40	0.36#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071822\
 Data File : VW023845.D
 Acq On : 18 Jul 2022 16:31
 Operator : SY/VA
 Sample : N3701-11
 Misc : 6.67g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 F5B84

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 07/19/2022
 Supervised By :Mahesh Dadoda 07/20/2022

Quant Time: Jul 19 01:09:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jul 19 01:05:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	213602	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.627	117	199366	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	85985	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	96575	17.702	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150	Recovery	=	70.800%	
7) Chloroethane-d5	2.873	69	82916	22.782	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150	Recovery	=	91.120%	
11) 1,1-Dichloroethene-d2	4.013	63	46234	10.407	ug/L	0.00
Spiked Amount 25.000	Range 45	- 110	Recovery	=	41.640%#	
21) 2-Butanone-d5	7.079	46	40561	47.932	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135	Recovery	=	95.860%	
24) Chloroform-d	7.653	84	124714	20.054	ug/L	0.00
Spiked Amount 25.000	Range 40	- 150	Recovery	=	80.200%	
26) 1,2-Dichloroethane-d4	8.305	65	75929	19.963	ug/L	0.00
Spiked Amount 25.000	Range 70	- 130	Recovery	=	79.840%	
32) Benzene-d6	8.274	84	208514	20.120	ug/L	0.00
Spiked Amount 25.000	Range 20	- 135	Recovery	=	80.480%	
36) 1,2-Dichloropropane-d6	9.274	67	71468	22.894	ug/L	0.00
Spiked Amount 25.000	Range 70	- 120	Recovery	=	91.560%	
41) Toluene-d8	10.323	98	193810	18.437	ug/L	0.00
Spiked Amount 25.000	Range 30	- 130	Recovery	=	73.760%	
43) trans-1,3-Dichloroprop...	10.573	79	26662	18.221	ug/L	0.00
Spiked Amount 25.000	Range 30	- 135	Recovery	=	72.880%	
47) 2-Hexanone-d5	10.926	63	32906	50.128	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135	Recovery	=	100.260%	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	75166	21.788	ug/L	0.00
Spiked Amount 25.000	Range 45	- 120	Recovery	=	87.160%	
66) 1,2-Dichlorobenzene-d4	13.853	152	60405	19.679	ug/L	0.00
Spiked Amount 25.000	Range 75	- 120	Recovery	=	78.720%	
Target Compounds						
13) Acetone	4.123	43	14881	24.873	ug/L	76
14) Carbon disulfide	4.379	76	15271	2.293	ug/L	97
16) Methylene chloride	4.915	84	13478	4.035	ug/L	93
53) m,p-Xylene	11.835	106	3578	0.625	ug/L	88
63) 1,2,4-Trimethylbenzene	13.249	105	7833m	0.828	ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071822\
Data File : VW023845.D
Acq On : 18 Jul 2022 16:31
Operator : SY/VA
Sample : N3701-11
Misc : 6.67g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
F5B84

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 07/19/2022
Supervised By :Mahesh Dadoda 07/20/2022

Quant Time: Jul 19 01:09:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM071322SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jul 19 01:05:58 2022
Response via : Initial Calibration

