

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071824\
 Data File : VW029633.D
 Acq On : 18 Jul 2024 15:28
 Operator : SY/MD
 Sample : P3264-10
 Misc : 5.75g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4BE3

Quant Time: Jul 19 01:15:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jul 19 01:07:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	473875	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	394211	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	148987	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	77243	10.090	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	40.360%
7) Chloroethane-d5	2.899	69	65479	11.879	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	47.520%
11) 1,1-Dichloroethene-d2	4.021	65	38753	11.273	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	45.080%
21) 2-Butanone-d5	7.075	46	32785	14.267	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	28.540%
24) Chloroform-d	7.648	84	172888	12.620	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	50.480%
26) 1,2-Dichloroethane-d4	8.300	65	97131	12.088	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	48.360%#
32) Benzene-d6	8.276	84	322942	13.675	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	54.720%
36) 1,2-Dichloropropane-d6	9.276	67	109071	14.445	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	57.800%#
41) Toluene-d8	10.324	98	271093	13.053	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	52.200%
43) trans-1,3-Dichloroprop...	10.574	79	33523	10.484	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	41.920%
47) 2-Hexanone-d5	10.922	63	22186	16.170	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	32.340%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	68395	11.156	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	44.640%#
66) 1,2-Dichlorobenzene-d4	13.848	152	68425	13.188	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	52.760%#
Target Compounds						
					Qvalue	
16) Methylene chloride	4.917	84	26254	3.095	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071824\
Data File : VW029633.D
Acq On : 18 Jul 2024 15:28
Operator : SY/MD
Sample : P3264-10
Misc : 5.75g/10mL/MSVOA_W/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4BE3

Quant Time: Jul 19 01:15:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jul 19 01:07:26 2024
Response via : Initial Calibration

