

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW071918\
 Data File : VW004063.D
 Acq On : 20 Jul 2018 04:17
 Operator : VA/AP
 Sample : J4023-14
 Misc : 7.39G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DB1L8

Quant Time: Jul 20 08:21:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM071818S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 20 07:22:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	314878	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	304735	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	133907	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	63722	20.46	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.84%
7) Chloroethane-d5	2.89	69	47531	21.17	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.68%
10) 1,1-Dichloroethene-d2	4.01	63	110959	14.85	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.40%
20) 2-Butanone-d5	7.08	46	65248	35.72	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	71.44%
24) Chloroform-d	7.65	84	183471	24.55	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.20%
26) 1,2-Dichloroethane-d4	8.31	65	110668	24.59	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	98.36%
29) Benzene-d6	8.28	84	313226	22.16	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	88.64%
33) 1,2-Dichloropropane-d6	9.28	67	115451	24.49	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	97.96%
37) Toluene-d8	10.33	98	270778	21.32	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.28%
38) trans-1,3-Dichloropropene-	10.59	79	43765	21.12	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.48%
39) 2-Hexanone-d5	10.93	63	43996	34.67	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	69.34%
48) 1,1,2,2-Tetrachloroethane-	12.70	84	114325	22.13	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	88.52%
61) 1,2-Dichlorobenzene-d4	13.87	152	113146	26.28	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	105.12%

Target Compounds					Qvalue
13) Acetone	4.12	43	10171	5.122 ug/L	98
16) Methylene chloride	4.90	84	13321	2.205 ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW071918\
Data File : VW004063.D
Acq On : 20 Jul 2018 04:17
Operator : VA/AP
Sample : J4023-14
Misc : 7.39G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DB1L8

Quant Time: Jul 20 08:21:45 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM071818S.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 20 07:22:04 2018
Response via : Initial Calibration

