

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071924\
 Data File : VW029663.D
 Acq On : 19 Jul 2024 17:57
 Operator : SY/MD
 Sample : P3264-08
 Misc : 5.09g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Jul 19 23:23:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jul 19 23:16:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	481301	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	403908	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	149387	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	92037	11.837	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	47.360%
7) Chloroethane-d5	2.899	69	77524	13.847	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	55.400%
11) 1,1-Dichloroethene-d2	4.015	65	44623	12.781	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	51.120%
21) 2-Butanone-d5	7.081	46	33938	14.541	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	29.080%
24) Chloroform-d	7.648	84	216563	15.564	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	62.240%
26) 1,2-Dichloroethane-d4	8.307	65	122832	15.051	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	60.200%#
32) Benzene-d6	8.276	84	387507	16.015	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	64.080%
36) 1,2-Dichloropropane-d6	9.270	67	135985	17.577	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	70.320%
41) Toluene-d8	10.325	98	318624	14.973	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	59.880%
43) trans-1,3-Dichloroprop...	10.575	79	37612	11.481	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	45.920%
47) 2-Hexanone-d5	10.922	63	20831	14.818	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	29.640%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	88930	14.158	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	56.640%
66) 1,2-Dichlorobenzene-d4	13.849	152	85946	16.520	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	66.080%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071924\
Data File : VW029663.D
Acq On : 19 Jul 2024 17:57
Operator : SY/MD
Sample : P3264-08
Misc : 5.09g/10mL/MSVOA_W/SOIL/B
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Jul 19 23:23:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jul 19 23:16:32 2024
Response via : Initial Calibration

