

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071924\
 Data File : VW029659.D
 Acq On : 19 Jul 2024 16:22
 Operator : SY/MD
 Sample : VIBLK499
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK499

Quant Time: Jul 19 23:22:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jul 19 23:16:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	241730	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	226155	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.550	152	99125	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	59665	15.279	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	61.120%		
7) Chloroethane-d5	2.899	69	49140	17.476	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	69.920%		
11) 1,1-Dichloroethene-d2	4.021	65	30190	17.217	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	68.880%		
21) 2-Butanone-d5	7.075	46	41667	35.545	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	71.100%		
24) Chloroform-d	7.648	84	134235	19.209	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	76.840%		
26) 1,2-Dichloroethane-d4	8.301	65	80264	19.582	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	78.320%		
32) Benzene-d6	8.276	84	253526	18.713	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	74.840%		
36) 1,2-Dichloropropane-d6	9.270	67	85233	19.676	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	78.720%		
41) Toluene-d8	10.319	98	216798	18.196	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	72.800%		
43) trans-1,3-Dichloroprop...	10.575	79	29942	16.323	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	65.280%		
47) 2-Hexanone-d5	10.922	63	26165	33.241	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	66.480%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	64388	18.308	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	73.240%		
66) 1,2-Dichlorobenzene-d4	13.849	152	72458	20.990	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	83.960%		
Target Compounds						Qvalue
16) Methylene chloride	4.917	84	21734	5.022	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW071924\
Data File : VW029659.D
Acq On : 19 Jul 2024 16:22
Operator : SY/MD
Sample : VIBLK499
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK499

Quant Time: Jul 19 23:22:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jul 19 23:16:32 2024
Response via : Initial Calibration

