

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072021\
 Data File : VW019506.D
 Acq On : 20 Jul 2021 10:04
 Operator : SY/VA
 Sample : VW0720SBL01
 Misc : 5.00g/5.0mL/MSVOA_W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0720SBL01

Quant Time: Jul 20 14:47:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 20 04:37:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.952	168	248541	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.848	114	435893	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.634	117	387092	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.560	152	169308	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.305	65	154852	65.831	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.660%
35) Dibromofluoromethane	7.891	113	138690	52.078	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.160%
50) Toluene-d8	10.323	98	539390	52.312	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.620%
62) 4-Bromofluorobenzene	12.621	95	178626	46.184	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.360%
Target Compounds						
16) Acetone	4.178	43	5601	12.391	ug/l	Qvalue # 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072021\
Data File : VW019506.D
Acq On : 20 Jul 2021 10:04
Operator : SY/VA
Sample : VW0720SBL01
Misc : 5.00g/5.0mL/MSVOA_W/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0720SBL01

Quant Time: Jul 20 14:47:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 20 04:37:55 2021
Response via : Initial Calibration

