

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072021\
 Data File : VW019511.D
 Acq On : 20 Jul 2021 15:08
 Operator : SY/VA
 Sample : M3072-01
 Misc : 5.04g/5.1mL/MSVOA_W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 OR-02-071621

Quant Time: Jul 20 16:05:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 20 04:37:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.958	168	12696	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.848	114	20767	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.634	117	19179	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.560	152	7710	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.311	65	8394	70.164	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	140.320%
35) Dibromofluoromethane	7.890	113	7306	57.583	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.160%
50) Toluene-d8	10.329	98	24825	50.535	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.080%
62) 4-Bromofluorobenzene	12.621	95	8868	48.126	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.260%
Target Compounds						
					Qvalue	
16) Acetone	4.135	43	6009	260.240	ug/l	96
20) Methylene Chloride	4.928	84	5326	33.627	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072021\
Data File : VW019511.D
Acq On : 20 Jul 2021 15:08
Operator : SY/VA
Sample : M3072-01
Misc : 5.04g/5.1mL/MSVOA_W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
OR-02-071621

Quant Time: Jul 20 16:05:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 20 04:37:55 2021
Response via : Initial Calibration

