

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072021\
 Data File : VW019514.D
 Acq On : 20 Jul 2021 16:24
 Operator : SY/VA
 Sample : M3039-09
 Misc : 5.08g/5.3mL/MSVOA_W/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-04-8.0-8.5

Quant Time: Jul 21 02:04:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 20 04:37:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.952	168	253217	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.848	114	474947	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.634	117	428561	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.560	152	205129	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.311	65	175020	73.579	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	147.160%
35) Dibromofluoromethane	7.885	113	152504	52.556	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.120%
50) Toluene-d8	10.323	98	591066	52.610	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.220%
62) 4-Bromofluorobenzene	12.621	95	209973	49.825	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.640%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.928	84	46436	11.856	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072021\
Data File : VW019514.D
Acq On : 20 Jul 2021 16:24
Operator : SY/VA
Sample : M3039-09
Misc : 5.08g/5.3mL/MSVOA_W/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-04-8.0-8.5

Quant Time: Jul 21 02:04:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 20 04:37:55 2021
Response via : Initial Calibration

