

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072021\
 Data File : VW019519.D
 Acq On : 20 Jul 2021 18:31
 Operator : SY/VA
 Sample : M3076-02
 Misc : 5.06g/5.3mL/MSVOA_W/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 20210587-AMENDED-COM-1

Quant Time: Jul 21 02:06:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 20 04:37:55 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.952	168	246523	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.842	114	472322	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.634	117	440736	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.560	152	199474	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.311	65	183708	79.721	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 159.440%	
35) Dibromofluoromethane	7.891	113	148288	51.387	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.780%	
50) Toluene-d8	10.323	98	602046	53.885	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.780%	
62) 4-Bromofluorobenzene	12.621	95	218111	52.043	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 104.080%	
Target Compounds						
					Qvalue	
16) Acetone	4.147	43	30773	68.636	ug/l	96
17) Carbon Disulfide	4.391	76	221820	28.618	ug/l	99
20) Methylene Chloride	4.928	84	32852	7.234	ug/l	97
25) 2-Butanone	7.196	43	6280	8.143	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072021\
Data File : VW019519.D
Acq On : 20 Jul 2021 18:31
Operator : SY/VA
Sample : M3076-02
Misc : 5.06g/5.3mL/MSVOA_W/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
20210587-AMENDED-COM-1

Quant Time: Jul 21 02:06:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 20 04:37:55 2021
Response via : Initial Calibration

