

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072118\
 Data File : VW004114.D
 Acq On : 21 Jul 2018 14:58
 Operator : SY/AP
 Sample : J4037-08
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK01

Quant Time: Jul 23 00:04:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM071818S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jul 21 06:22:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	327924	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	318073	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	144994	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	69161	21.32	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.28%
7) Chloroethane-d5	2.89	69	51225	21.91	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.64%
10) 1,1-Dichloroethene-d2	4.00	63	126250	16.22	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.88%
20) 2-Butanone-d5	7.08	46	94773	49.83	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.66%
24) Chloroform-d	7.65	84	200938	25.82	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	103.28%
26) 1,2-Dichloroethane-d4	8.31	65	122443	26.12	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	104.48%
29) Benzene-d6	8.28	84	354321	24.02	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.08%
33) 1,2-Dichloropropane-d6	9.27	67	127375	25.88	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	103.52%
37) Toluene-d8	10.33	98	307667	23.21	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.84%
38) trans-1,3-Dichloropropene-	10.58	79	49706	22.99	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.96%
39) 2-Hexanone-d5	10.93	63	64326	48.56	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.12%
48) 1,1,2,2-Tetrachloroethane-	12.70	84	137223	25.45	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	101.80%
61) 1,2-Dichlorobenzene-d4	13.87	152	130687	28.03	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	112.12%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072118\
Data File : VW004114.D
Acq On : 21 Jul 2018 14:58
Operator : SY/AP
Sample : J4037-08
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK01

Quant Time: Jul 23 00:04:53 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM071818S.M
Quant Title : VOC Analysis
QLast Update : Sat Jul 21 06:22:08 2018
Response via : Initial Calibration

