

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072318\
 Data File : VW004154.D
 Acq On : 24 Jul 2018 00:17
 Operator : SY/AP
 Sample : VW0723SBL02
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0723SBL02

Quant Time: Jul 24 06:41:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 16 15:06:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	213108	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	374203	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	354553	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	167780	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	150677	63.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	126.50%	
35) Dibromofluoromethane	7.88	113	127308	53.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.04%	
50) Toluene-d8	10.33	98	461941	49.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.16%	
62) 4-Bromofluorobenzene	12.62	95	153993	45.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.68%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW072318\
Data File : VW004154.D
Acq On : 24 Jul 2018 00:17
Operator : SY/AP
Sample : VW0723SBL02
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0723SBL02

Quant Time: Jul 24 06:41:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 16 15:06:12 2018
Response via : Initial Calibration

