

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072318\
 Data File : VW004157.D
 Acq On : 24 Jul 2018 01:35
 Operator : SY/AP
 Sample : J4062-01RE
 Misc : 6.35G/5ML/MSVOA W/SOIL
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-3RE

Quant Time: Jul 24 06:48:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 16 15:06:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	62014	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	132838	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	122912	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	36618	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	73475	106.00	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	212.00%	
35) Dibromofluoromethane	7.88	113	57756	67.76	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	135.52%	
50) Toluene-d8	10.33	98	132780	40.15	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	80.30%	
62) 4-Bromofluorobenzene	12.62	95	37983	31.85	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	63.70%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.12	43	24262	145.65	ug/l	95
17) Carbon Disulfide	4.37	76	7632	4.06	ug/l #	92
18) Methyl Acetate	4.67	43	5408	13.53	ug/l	100
20) Methylene Chloride	4.91	84	13157	11.90	ug/l	94
30) Chloroform	7.67	83	3119	2.56	ug/l #	71
52) Toluene	10.38	92	2981	1.27	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW072318\
Data File : VW004157.D
Acq On : 24 Jul 2018 01:35
Operator : SY/AP
Sample : J4062-01RE
Misc : 6.35G/5ML/MSVOA W/SOIL
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-3RE

Quant Time: Jul 24 06:48:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 16 15:06:12 2018
Response via : Initial Calibration

