

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072318\  
 Data File : VW004161.D  
 Acq On : 24 Jul 2018 03:19  
 Operator : SY/AP  
 Sample : J4062-25  
 Misc : 9.99G/5ML/MSVOA W/SOIL  
 ALS Vial : 39 Sample Multiplier: 1

Instrument :  
 MSVOA\_W  
 ClientSampleID :  
 TP-9

Quant Time: Jul 24 06:59:57 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_W\METHOD\82W071618S.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jul 16 15:06:12 2018  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	23054	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	42154	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	37958	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	10249	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	26150	101.48	ug/l	0.00
Spiked Amount				50.000		
Recovery						202.96%
35) Dibromofluoromethane	7.89	113	19981	73.88	ug/l	0.00
Spiked Amount				50.000		
Recovery						147.76%
50) Toluene-d8	10.33	98	37041	35.29	ug/l	0.00
Spiked Amount				50.000		
Recovery						70.58%
62) 4-Bromofluorobenzene	12.62	95	10655	28.15	ug/l	0.00
Spiked Amount				50.000		
Recovery						56.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.12	43	2872	46.38	ug/l	91
20) Methylene Chloride	4.90	84	4673	10.95	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072318\  
 Data File : VW004161.D  
 Acq On : 24 Jul 2018 03:19  
 Operator : SY/AP  
 Sample : J4062-25  
 Misc : 9.99G/5ML/MSVOA W/SOIL  
 ALS Vial : 39 Sample Multiplier: 1

Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 TP-9

Quant Time: Jul 24 06:59:57 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_W\METHOD\82W071618S.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jul 16 15:06:12 2018  
 Response via : Initial Calibration

