

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072318\
 Data File : VW004164.D
 Acq On : 24 Jul 2018 04:37
 Operator : SY/AP
 Sample : J4036-03
 Misc : 8.20G/5ML/MSVOA W/SOIL
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 CON-SB-03-0-2

Quant Time: Jul 24 07:22:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 16 15:06:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	129181	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	215666	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	139676	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	65680	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	92969	64.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	128.76%	
35) Dibromofluoromethane	7.88	113	78366	56.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.26%	
50) Toluene-d8	10.33	98	95920	17.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	35.72%	
62) 4-Bromofluorobenzene	12.62	95	56405	29.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	58.26%	
Target Compounds						
16) Acetone	4.13	43	6593	19.00	ug/l	Qvalue 91
30) Chloroform	7.68	83	3374	1.33	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW072318\
Data File : VW004164.D
Acq On : 24 Jul 2018 04:37
Operator : SY/AP
Sample : J4036-03
Misc : 8.20G/5ML/MSVOA W/SOIL
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
CON-SB-03-0-2

Quant Time: Jul 24 07:22:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 16 15:06:12 2018
Response via : Initial Calibration

