


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW072318\
 Data File : VW004152.D
 Acq On : 23 Jul 2018 22:33
 Operator : SY/AP
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
 Title : SW846 8260
 Last Update : Mon Jul 16 15:06:12 2018



AutoFind: Scans 1810, 1811, 1812; Background Corrected with Scan 1801

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.3	15042	PASS
75	95	30	60	51.9	38450	PASS
95	95	100	100	100.0	74146	PASS
96	95	5	9	6.8	5008	PASS
173	174	0.00	2	0.7	411	PASS
174	95	50	100	83.4	61858	PASS
175	174	5	9	7.9	4910	PASS
176	174	95	101	96.3	59541	PASS
177	176	5	9	6.3	3743	PASS