

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072321\
 Data File : VW019574.D
 Acq On : 23 Jul 2021 15:35
 Operator : SY/VA
 Sample : M3138-01
 Misc : 5.05g/5.0mL/MSVOA_W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VB16085

Quant Time: Jul 24 05:19:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 20 04:37:55 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.952	168	34548	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.848	114	52345	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.634	117	45264	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.554	152	18384	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.311	65	18562	56.078	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.160%
35) Dibromofluoromethane	7.884	113	17548	54.871	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.740%
50) Toluene-d8	10.323	98	63703	51.447	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.900%
62) 4-Bromofluorobenzene	12.621	95	19086	41.093	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	82.180%
Target Compounds						Qvalue
16) Acetone	4.135	43	3207	51.040	ug/l	# 86
20) Methylene Chloride	4.915	84	10858	23.926	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072321\
Data File : VW019574.D
Acq On : 23 Jul 2021 15:35
Operator : SY/VA
Sample : M3138-01
Misc : 5.05g/5.0mL/MSVOA_W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VB16085

Quant Time: Jul 24 05:19:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 20 04:37:55 2021
Response via : Initial Calibration

