

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072321\
 Data File : VW019577.D
 Acq On : 23 Jul 2021 16:51
 Operator : SY/VA
 Sample : M3150-04
 Misc : 5.21g/5.0mL/MSVOA_W/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-C

Quant Time: Jul 24 05:20:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 20 04:37:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.951	168	264047	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.841	114	487215	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.634	117	442705	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.560	152	201350	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.305	65	169827	68.119	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	136.240%
35) Dibromofluoromethane	7.890	113	153714	51.639	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.280%
50) Toluene-d8	10.329	98	608885	52.831	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.660%
62) 4-Bromofluorobenzene	12.621	95	213089	49.291	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.580%
Target Compounds						
					Qvalue	
16) Acetone	4.153	43	56699	118.068	ug/l	100
20) Methylene Chloride	4.921	84	46394	11.148	ug/l	98
25) 2-Butanone	7.183	43	16962	20.533	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072321\
Data File : VW019577.D
Acq On : 23 Jul 2021 16:51
Operator : SY/VA
Sample : M3150-04
Misc : 5.21g/5.0mL/MSVOA_W/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-C

Quant Time: Jul 24 05:20:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 20 04:37:55 2021
Response via : Initial Calibration

