

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072321\
 Data File : VW019578.D
 Acq On : 23 Jul 2021 17:17
 Operator : SY/VA
 Sample : M3143-02
 Misc : 6.18g/5.0mL/MSVOA_W/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 STOCK-PILE-1

Quant Time: Jul 24 05:20:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 20 04:37:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.951	168	220264	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.848	114	410891	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.634	117	371578	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.560	152	159140	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.311	65	157643	76.367	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	152.740%
35) Dibromofluoromethane	7.890	113	132592	52.818	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.640%
50) Toluene-d8	10.323	98	506297	52.090	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.180%
62) 4-Bromofluorobenzene	12.621	95	177511	48.688	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.380%
Target Compounds						
					Qvalue	
16) Acetone	4.153	43	78726	196.523	ug/l	98
25) 2-Butanone	7.183	43	29948	43.459	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072321\
Data File : VW019578.D
Acq On : 23 Jul 2021 17:17
Operator : SY/VA
Sample : M3143-02
Misc : 6.18g/5.0mL/MSVOA_W/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
STOCK-PILE-1

Quant Time: Jul 24 05:20:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 20 04:37:55 2021
Response via : Initial Calibration

