

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072321\
 Data File : VW019582.D
 Acq On : 23 Jul 2021 18:58
 Operator : SY/VA
 Sample : M3151-03
 Misc : 5.48g/5.0mL/MSVOA_W/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 COMP-3

Quant Time: Jul 24 05:21:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 20 04:37:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.952	168	247906	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.848	114	466949	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.634	117	423799	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.554	152	184911	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.311	65	160136	68.436	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 136.880%	
35) Dibromofluoromethane	7.891	113	149856	52.528	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.060%	
50) Toluene-d8	10.323	98	587772	53.213	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 106.420%	
62) 4-Bromofluorobenzene	12.621	95	196116	47.333	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 94.660%	
Target Compounds						
					Qvalue	
16) Acetone	4.135	43	8802	19.522	ug/l	98
20) Methylene Chloride	4.922	84	66848	19.810	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072321\
Data File : VW019582.D
Acq On : 23 Jul 2021 18:58
Operator : SY/VA
Sample : M3151-03
Misc : 5.48g/5.0mL/MSVOA_W/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
COMP-3

Quant Time: Jul 24 05:21:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 20 04:37:55 2021
Response via : Initial Calibration

