

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072321\
 Data File : VW019583.D
 Acq On : 23 Jul 2021 19:23
 Operator : SY/VA
 Sample : M3129-01
 Misc : 5.05g/5.0mL/MSVOA_W/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 RO-VNJ-207

Quant Time: Jul 24 05:21:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 20 04:37:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.951	168	151	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.860	114	230	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.640	117	62	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.56
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.317	65	43	27.364	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	54.720%
35) Dibromofluoromethane	7.884	113	108	76.857	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	153.720%#
50) Toluene-d8	10.335	98	282	51.832	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.660%
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	25 - 144	Recovery	=	0.000%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 24 05:21:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 20 04:37:55 2021
Response via : Initial Calibration

