

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072321\
 Data File : VW019585.D
 Acq On : 23 Jul 2021 20:14
 Operator : SY/VA
 Sample : M3152-01
 Misc : 6.61g/5.0mL/MSVOA_W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VAULT-A

Quant Time: Jul 24 05:21:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 20 04:37:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.951	168	232561	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.848	114	450501	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.634	117	409161	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.560	152	183559	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.305	65	167497	76.882	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 153.760%	
35) Dibromofluoromethane	7.891	113	149222	54.216	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.440%	
50) Toluene-d8	10.323	98	566526	53.162	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 106.320%	
62) 4-Bromofluorobenzene	12.621	95	200314	50.112	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 100.220%	
Target Compounds						
						Qvalue
16) Acetone	4.135	43	5140	12.152	ug/l	88
52) Toluene	10.390	92	11516	1.380	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072321\
Data File : VW019585.D
Acq On : 23 Jul 2021 20:14
Operator : SY/VA
Sample : M3152-01
Misc : 6.61g/5.0mL/MSVOA_W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VAULT-A

Quant Time: Jul 24 05:21:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 20 04:37:55 2021
Response via : Initial Calibration

