

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072321\
 Data File : VW019581.D
 Acq On : 23 Jul 2021 18:33
 Operator : SY/VA
 Sample : M3151-02
 Misc : 4.88g/5.0mL/MSVOA_W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 COMP-2

Quant Time: Jul 24 05:21:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 20 04:37:55 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.951	168	243426	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.848	114	460033	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.634	117	412153	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.560	152	181405	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.311	65	158864	69.194	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 138.380%	
35) Dibromofluoromethane	7.890	113	147096	52.336	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.680%	
50) Toluene-d8	10.323	98	586206	53.869	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.740%	
62) 4-Bromofluorobenzene	12.621	95	194497	47.648	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 95.300%	
Target Compounds						
						Qvalue
16) Acetone	4.147	43	33797	76.340	ug/l	97
20) Methylene Chloride	4.915	84	60464	17.849	ug/l	95
25) 2-Butanone	7.183	43	11213	14.724	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072321\
Data File : VW019581.D
Acq On : 23 Jul 2021 18:33
Operator : SY/VA
Sample : M3151-02
Misc : 4.88g/5.0mL/MSVOA_W/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
COMP-2

Quant Time: Jul 24 05:21:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W071921S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 20 04:37:55 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.951 min Scan# 992
Delta R.T. 0.000 min
Lab File: VW019581.D
Acq: 23 Jul 2021 18:33




Instrument :
MSVOA_W
ClientSampleId :
COMP-2


Tgt Ion:168 Resp: 243426
Ion Ratio Lower Upper
168 100
99 58.4 51.9 77.9



#16
Acetone
Concen: 76.340 ug/l
RT: 4.147 min Scan# 368
Delta R.T. 0.006 min
Lab File: VW019581.D
Acq: 23 Jul 2021 18:33

Tgt Ion: 43 Resp: 33797
Ion Ratio Lower Upper
43 100
58 32.4 27.2 40.8









#20
Methylene Chloride
Concen: 17.849 ug/l
RT: 4.915 min Scan# 494
Delta R.T. -0.007 min
Lab File: VW019581.D
Acq: 23 Jul 2021 18:33




Instrument :
MSVOA_W
ClientSampleId :
COMP-2


Tgt Ion:	84	Resp:	60464
Ion	Ratio	Lower	Upper
84	100		
49	114.3	97.2	145.8
51	34.0	30.1	45.1
86	60.5	49.9	74.9



#25
2-Butanone
Concen: 14.724 ug/l
RT: 7.183 min Scan# 866
Delta R.T. 0.006 min
Lab File: VW019581.D
Acq: 23 Jul 2021 18:33

Tgt Ion:	43	Resp:	11213
Ion	Ratio	Lower	Upper
43	100		
72	27.6	21.4	32.0









#33
1,2-Dichloroethane-d4
Concen: 69.194 ug/l
RT: 8.311 min Scan# 1051
Delta R.T. 0.000 min
Lab File: VW019581.D
Acq: 23 Jul 2021 18:33




Instrument :
MSVOA_W
ClientSampled :
COMP-2


Tgt Ion: 65 Resp: 158864
Ion Ratio Lower Upper
65 100
67 53.0 0.0 107.8



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.848 min Scan# 1139
Delta R.T. 0.000 min
Lab File: VW019581.D
Acq: 23 Jul 2021 18:33

Tgt Ion: 114 Resp: 460033
Ion Ratio Lower Upper
114 100
63 21.1 0.0 43.8
88 17.1 0.0 33.4









#35
Dibromofluoromethane
Concen: 52.336 ug/l
RT: 7.890 min Scan# 982
Delta R.T. 0.000 min
Lab File: VW019581.D
Acq: 23 Jul 2021 18:33




Instrument :
MSVOA_W
ClientSampleId :
COMP-2


Tgt Ion:113 Resp: 147096
Ion Ratio Lower Upper
113 100
111 104.0 83.8 125.6
192 19.3 15.4 23.2



#50
Toluene-d8
Concen: 53.869 ug/l
RT: 10.323 min Scan# 1381
Delta R.T. -0.006 min
Lab File: VW019581.D
Acq: 23 Jul 2021 18:33

Tgt Ion: 98 Resp: 586206
Ion Ratio Lower Upper
98 100
100 64.1 52.2 78.4









#62
4-Bromofluorobenzene
Concen: 47.648 ug/l
RT: 12.621 min Scan# 1758
Delta R.T. 0.000 min
Lab File: VW019581.D
Acq: 23 Jul 2021 18:33

Instrument :
MSVOA_W
ClientSampleId :
COMP-2

Tgt Ion: 95 Resp: 194497




Ion	Ratio	Lower	Upper
95	100		
174	74.0	0.0	146.6
176	72.3	0.0	136.8




#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.634 min Scan# 1596
Delta R.T. 0.000 min
Lab File: VW019581.D
Acq: 23 Jul 2021 18:33

Tgt Ion: 117 Resp: 412153

Ion	Ratio	Lower	Upper
117	100		
82	56.7	47.4	71.0
119	31.0	26.6	39.8





#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.560 min Scan# 1912
Delta R.T. 0.000 min
Lab File: VW019581.D
Acq: 23 Jul 2021 18:33

Instrument :
MSVOA_W
ClientSampleId :
COMP-2

Tgt Ion:152 Resp: 181405
Ion Ratio Lower Upper
152 100
115 61.3 30.6 91.8
150 158.3 0.0 342.4

