

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072419\
 Data File : VW011460.D
 Acq On : 24 Jul 2019 04:24
 Operator : SY/VA
 Sample : K3934-10
 Misc : 5.85G/5ML/MSVOA W/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-1S-2

Quant Time: Jul 24 08:06:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070819S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 11 08:14:17 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	247934	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	413564	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	327436	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	156711	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	154359	49.55	ug/l	0.00
Spiked Amount			Recovery	=	99.10%	
35) Dibromofluoromethane	7.88	113	122522	49.70	ug/l	0.00
Spiked Amount			Recovery	=	99.40%	
50) Toluene-d8	10.32	98	509857	51.17	ug/l	0.00
Spiked Amount			Recovery	=	102.34%	
62) 4-Bromofluorobenzene	12.62	95	202724	55.51	ug/l	0.00
Spiked Amount			Recovery	=	111.02%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.13	43	75358	100.531	ug/l	97
17) Carbon Disulfide	4.38	76	12951	1.551	ug/l	97
25) 2-Butanone	7.18	43	26411	22.446	ug/l	97
40) Benzene	8.32	78	252777	20.769	ug/l	99
52) Toluene	10.39	92	29197	3.955	ug/l	96
67) Ethyl Benzene	11.73	91	32914	2.583	ug/l	97
68) m/p-Xylenes	11.83	106	13432	2.913	ug/l	87
69) o-Xylene	12.16	106	8142	1.873	ug/l	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW072419\
Data File : VW011460.D
Acq On : 24 Jul 2019 04:24
Operator : SY/VA
Sample : K3934-10
Misc : 5.85G/5ML/MSVOA W/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-1S-2

Quant Time: Jul 24 08:06:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070819S.M
Quant Title : SW846 8260
QLast Update : Thu Jul 11 08:14:17 2019
Response via : Initial Calibration

