

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072519\
 Data File : VW011486.D
 Acq On : 24 Jul 2019 20:10
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 25 05:52:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070819S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 11 08:14:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.373	0.273	26.8#	75	0.00
3 P	Chloromethane	0.481	0.430	10.6	93	0.00
4 C	Vinyl Chloride	0.629	0.548	12.9#	86	0.00
5 T	Bromomethane	0.316	0.271	14.2	84	-0.01
6 T	Chloroethane	0.339	0.307	9.4	87	0.00
7 T	Trichlorofluoromethane	0.223	0.236	-5.8	103	0.00
8 T	Diethyl Ether	0.309	0.280	9.4	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.535	0.475	11.2	86	0.00
10 T	Methyl Iodide	0.736	0.685	6.9	89	0.00
11 T	Tert butyl alcohol	0.047	0.043	8.5	93	0.00
12 CM	1,1-Dichloroethene	0.555	0.504	9.2#	88	0.00
13 T	Acrolein	0.037	0.034	8.1	91	0.00
14 T	Allyl chloride	1.013	1.011	0.2	94	0.00
15 T	Acrylonitrile	0.157	0.151	3.8	94	0.00
16 T	Acetone	0.151	0.153	-1.3	99	0.00
17 T	Carbon Disulfide	1.684	1.411	16.2	80	0.00
18 T	Methyl Acetate	0.473	0.396	16.3	102	0.00
19 T	Methyl tert-butyl Ether	0.843	0.808	4.2	93	0.00
20 T	Methylene Chloride	0.596	0.582	2.3	97	0.00
21 T	trans-1,2-Dichloroethene	0.577	0.530	8.1	88	0.00
22 T	Diisopropyl ether	2.179	2.028	6.9	89	0.00
23 T	Vinyl Acetate	1.348	1.294	4.0	88	0.00
24 P	1,1-Dichloroethane	1.144	1.068	6.6	90	0.00
25 T	2-Butanone	0.237	0.227	4.2	95	0.00
26 T	2,2-Dichloropropane	0.595	0.559	6.1	90	0.00
27 T	cis-1,2-Dichloroethene	0.623	0.585	6.1	91	0.00
28 T	Bromochloromethane	0.520	0.446	14.2	86	0.00
29 T	Tetrahydrofuran	0.156	0.144	7.7	92	0.00
30 C	Chloroform	1.059	0.975	7.9#	91	0.00
31 T	Cyclohexane	1.217	1.039	14.6	86	0.00
32 T	1,1,1-Trichloroethane	0.806	0.744	7.7	88	0.00
33 S	1,2-Dichloroethane-d4	0.628	0.564	10.2	91	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.298	0.285	4.4	94	0.00
36 T	1,1-Dichloropropene	0.532	0.493	7.3	87	0.00
37 T	Ethyl Acetate	0.320	0.297	7.2	90	0.00
38 T	Carbon Tetrachloride	0.451	0.419	7.1	87	0.00
39 T	Methylcyclohexane	0.645	0.596	7.6	85	0.00
40 TM	Benzene	1.471	1.391	5.4	89	0.00
41 T	Methacrylonitrile	0.183	0.188	-2.7	104	0.00
42 TM	1,2-Dichloroethane	0.477	0.435	8.8	87	0.00
43 T	Isopropyl Acetate	0.604	0.563	6.8	89	0.00
44 TM	Trichloroethene	0.354	0.338	4.5	91	0.00
45 C	1,2-Dichloropropane	0.398	0.379	4.8#	90	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072519\
 Data File : VW011486.D
 Acq On : 24 Jul 2019 20:10
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 25 05:52:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070819S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 11 08:14:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.187	0.176	5.9	90	0.00
47 T	Bromodichloromethane	0.479	0.446	6.9	88	0.00
48 T	Methyl methacrylate	0.291	0.274	5.8	89	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	90	0.00
50 S	Toluene-d8	1.205	1.153	4.3	93	0.00
51 T	4-Methyl-2-Pentanone	0.316	0.301	4.7	92	0.00
52 CM	Toluene	0.893	0.854	4.4#	89	0.00
53 T	t-1,3-Dichloropropene	0.510	0.462	9.4	85	0.00
54 T	cis-1,3-Dichloropropene	0.600	0.553	7.8	86	0.00
55 T	1,1,2-Trichloroethane	0.266	0.250	6.0	91	0.00
56 T	Ethyl methacrylate	0.404	0.382	5.4	90	0.00
57 T	1,3-Dichloropropane	0.502	0.465	7.4	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.187	0.184	1.6	97	0.00
59 T	2-Hexanone	0.219	0.208	5.0	91	0.00
60 T	Dibromochloromethane	0.291	0.272	6.5	88	0.00
61 T	1,2-Dibromoethane	0.252	0.235	6.7	89	0.00
62 S	4-Bromofluorobenzene	0.442	0.423	4.3	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.337	0.320	5.0	91	0.00
65 PM	Chlorobenzene	1.017	0.967	4.9	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.358	0.342	4.5	90	0.00
67 C	Ethyl Benzene	1.946	1.871	3.9#	89	0.00
68 T	m/p-Xylenes	0.704	0.686	2.6	91	0.00
69 T	o-Xylene	0.664	0.639	3.8	90	0.00
70 T	Styrene	1.145	1.127	1.6	91	0.00
71 P	Bromoform	0.190	0.178	6.3	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	3.738	3.648	2.4	90	0.00
74 T	N-amyl acetate	1.302	1.205	7.5	88	0.00
75 P	1,1,2,2-Tetrachloroethane	0.766	0.707	7.7	89	0.00
76 T	1,2,3-Trichloropropane	0.578	0.503	13.0	91	0.00
77 T	Bromobenzene	0.839	0.798	4.9	91	0.00
78 T	n-propylbenzene	4.604	4.434	3.7	89	0.00
79 T	2-Chlorotoluene	2.650	2.547	3.9	90	0.00
80 T	1,3,5-Trimethylbenzene	3.175	3.048	4.0	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.255	0.228	10.6	85	0.00
82 T	4-Chlorotoluene	2.760	2.656	3.8	91	0.00
83 T	tert-Butylbenzene	2.648	2.587	2.3	91	0.00
84 T	1,2,4-Trimethylbenzene	3.188	3.053	4.2	90	0.00
85 T	sec-Butylbenzene	3.821	3.659	4.2	88	0.00
86 T	p-Isopropyltoluene	3.427	3.316	3.2	89	0.00
87 T	1,3-Dichlorobenzene	1.614	1.539	4.6	91	0.00
88 T	1,4-Dichlorobenzene	1.625	1.531	5.8	89	0.00
89 T	n-Butylbenzene	3.419	3.307	3.3	88	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072519\
Data File : VW011486.D
Acq On : 24 Jul 2019 20:10
Operator : SY/VA
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Jul 25 05:52:55 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070819S.M
Quant Title : SW846 8260
QLast Update : Thu Jul 11 08:14:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.626	0.578	7.7	86	0.00
91 T	1,2-Dichlorobenzene	1.464	1.378	5.9	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.136	0.122	10.3	87	0.00
93 T	1,2,4-Trichlorobenzene	1.008	1.003	0.5	92	0.00
94 T	Hexachlorobutadiene	0.619	0.610	1.5	90	0.00
95 T	Naphthalene	1.932	1.912	1.0	88	0.00
96 T	1,2,3-Trichlorobenzene	0.890	0.881	1.0	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6