

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072519\
 Data File : VW011476.D
 Acq On : 24 Jul 2019 15:47
 Operator : SY/VA
 Sample : K3880-03
 Misc : 17.33G/5ML/MSVOA W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SUP-5B-G-3-071719

Quant Time: Jul 25 04:59:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W070819S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 11 08:14:17 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	265635	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	445480	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	375074	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	151973	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	175670	52.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.26%	
35) Dibromofluoromethane	7.88	113	135635	51.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.16%	
50) Toluene-d8	10.32	98	544579	50.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.48%	
62) 4-Bromofluorobenzene	12.62	95	176793	44.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.88%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW072519\
Data File : VW011476.D
Acq On : 24 Jul 2019 15:47
Operator : SY/VA
Sample : K3880-03
Misc : 17.33G/5ML/MSVOA W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SUP-5B-G-3-071719

Quant Time: Jul 25 04:59:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W070819S.M
Quant Title : SW846 8260
QLast Update : Thu Jul 11 08:14:17 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.95 min Scan# 991

Delta R.T. 0.00 min

Lab File: VW011476.D

Acq: 24 Jul 2019 15:47

Instrument :

MSVOA_W

ClientSampleId :


SUP-5B-G-3-071719

Tot Ion:168 Resp: 265635

Ion Ratio Lower Upper

168 100

99 59.0 50.1 75.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VW



7.95

100000

50000

0

Time--> 7.80 7.90 8.00 8.10



#33

1,2-Dichloroethane-d4

Concen: 52.630 ug/l

RT: 8.31 min Scan# 1050

Delta R.T. 0.00 min

Lab File: VW011476.D


Acq: 24 Jul 2019 15:47

Tgt Ion: 65 Resp: 175670

Ion Ratio Lower Upper

65 100

67 52.6 0.0 104.4



Abundance Ion 64.90 (64.60 to 65.60): VW

Ion 66.90 (66.60 to 67.60): VW

8.31

80000


60000


40000

20000

0

Time--> 8.20 8.30 8.40









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.84 min Scan# 1138
 Delta R.T. 0.00 min
 Lab File: VW011476.D
 Acq: 24 Jul 2019 15:47

Instrument :
 MSVOA_W
 ClientSampleId :
 SUP-5B-G-3-071719

Tot Ion	Ratio	Lower	Upper
114	100		
63	24.7	0.0	50.8
88	17.0	0.0	34.2




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VW
 Ion 87.90 (87.60 to 88.60): VW






#35
 Dibromofluoromethane
 Concen: 51.082 ug/l
 RT: 7.88 min Scan# 981
 Delta R.T. 0.00 min
 Lab File: VW011476.D
 Acq: 24 Jul 2019 15:47

Tgt Ion	Ratio	Lower	Upper
113	100		
111	103.7	82.4	123.6
192	21.0	15.8	23.8



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 50.739 ug/l

RT: 10.32 min Scan# 1381

Delta R.T. 0.00 min

Lab File: VW011476.D

Acq: 24 Jul 2019 15:47

Instrument :

MSVOA_W

ClientSampleId :


SUP-5B-G-3-071719

Tot Ion: 98 Resp: 544579

Ion Ratio Lower Upper




98 100

100 65.5 52.4 78.6



Abundance Ion 98.00 (97.70 to 98.70): VW

Ion 100.00 (99.70 to 100.70): VW



#62

4-Bromofluorobenzene

Concen: 44.940 ug/l

RT: 12.62 min Scan# 1757

Delta R.T. -0.01 min

Lab File: VW011476.D

Acq: 24 Jul 2019 15:47


Tgt Ion: 95 Resp: 176793

Ion Ratio Lower Upper

95 100

174 77.0 0.0 147.6



176 73.9 0.0 144.6




Abundance Ion 94.90 (94.60 to 95.60): VW

Ion 173.80 (173.50 to 174.50): VW

Ion 175.80 (175.50 to 176.50): VW









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.63 min Scan# 1596
Delta R.T. 0.01 min
Lab File: VW011476.D
Acq: 24 Jul 2019 15:47

Instrument :
MSVOA_W
ClientSampleId :
SUP-5B-G-3-071719

Tot Ion:117 Resp: 375074
Ion Ratio Lower Upper
117 100
82 59.7 48.9 73.3
119 31.2 25.0 37.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.56 min Scan# 1912
Delta R.T. 0.00 min
Lab File: VW011476.D
Acq: 24 Jul 2019 15:47

Tgt Ion:152 Resp: 151973
Ion Ratio Lower Upper
152 100
115 60.7 30.1 90.3
150 158.9 0.0 346.0

