

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072522\
 Data File : VW024071.D
 Acq On : 25 Jul 2022 16:31
 Operator : SY/VA
 Sample : N3823-01
 Misc : 7.80g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 H0B60

Quant Time: Jul 26 01:33:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072522SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jul 26 01:30:33 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		8.842	114	243352	25.000	ug/L	0.00
28) Chlorobenzene-d5		11.628	117	239927	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		13.554	152	111376	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		2.343	65	131382	25.187	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	100.760%	
7) Chloroethane-d5		2.873	69	89741	25.462	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.840%	
11) 1,1-Dichloroethene-d2		4.007	63	77694	16.916	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.680%	
21) 2-Butanone-d5		7.080	46	69540	75.012	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	150.020%#	
24) Chloroform-d		7.647	84	163397	24.423	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.680%	
26) 1,2-Dichloroethane-d4		8.305	65	108075	28.117	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	112.480%	
32) Benzene-d6		8.275	84	294537	26.427	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	105.720%	
36) 1,2-Dichloropropane-d6		9.274	67	94926	27.003	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	108.000%	
41) Toluene-d8		10.323	98	274418	24.620	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	98.480%	
43) trans-1,3-Dichloroprop...		10.579	79	43104	27.654	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	110.600%	
47) 2-Hexanone-d5		10.920	63	58974	80.244	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	160.480%#	
56) 1,1,2,2-Tetrachloroeth...		12.688	84	118918	29.317	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	117.280%	
66) 1,2-Dichlorobenzene-d4		13.853	152	104961	29.099	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	116.400%	
Target Compounds							
13) Acetone		4.135	43	3655	5.556	ug/L	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072522\
Data File : VW024071.D
Acq On : 25 Jul 2022 16:31
Operator : SY/VA
Sample : N3823-01
Misc : 7.80g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
H0B60

Quant Time: Jul 26 01:33:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072522SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jul 26 01:30:33 2022
Response via : Initial Calibration

