

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072522\
 Data File : VW024073.D
 Acq On : 25 Jul 2022 17:20
 Operator : SY/VA
 Sample : N3823-03
 Misc : 7.36g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 H0B62

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 07/26/2022
 Supervised By :Mahesh Dadoda 07/26/2022

Quant Time: Jul 26 01:33:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072522SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jul 26 01:30:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	228355	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	235895	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	113272	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.337	65	139802	28.561	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	114.240%	
7) Chloroethane-d5	2.873	69	97200	29.390	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	117.560%	
11) 1,1-Dichloroethene-d2	4.007	63	78111	18.123	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	72.480%	
21) 2-Butanone-d5	7.080	46	60708	69.786	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	139.580%#	
24) Chloroform-d	7.647	84	164276	26.167	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	104.680%	
26) 1,2-Dichloroethane-d4	8.299	65	104463m	28.963	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	115.840%	
32) Benzene-d6	8.275	84	297996	27.194	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	108.760%	
36) 1,2-Dichloropropane-d6	9.274	67	98208	28.414	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	113.640%	
41) Toluene-d8	10.323	98	279988	25.549	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	102.200%	
43) trans-1,3-Dichloroprop...	10.579	79	41066	26.797	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	107.200%	
47) 2-Hexanone-d5	10.927	63	51475	71.237	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	142.480%#	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	109237	27.391	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	109.560%	
66) 1,2-Dichlorobenzene-d4	13.853	152	108607	29.606	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	118.440%	
Target Compounds					Qvalue	
13) Acetone	4.129	43	3282m	5.317	ug/L	
16) Methylene chloride	4.915	84	14688	4.226	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072522\
Data File : VW024073.D
Acq On : 25 Jul 2022 17:20
Operator : SY/VA
Sample : N3823-03
Misc : 7.36g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
H0B62

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 07/26/2022
Supervised By :Mahesh Dadoda 07/26/2022

Quant Time: Jul 26 01:33:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM072522SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jul 26 01:30:33 2022
Response via : Initial Calibration

