

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW072618\
 Data File : VW004209.D
 Acq On : 25 Jul 2018 22:55
 Operator : SY/AP
 Sample : J4123-01
 Misc : 6.94G/5ML/MSVOA_W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 RO-1-3

Quant Time: Jul 26 05:13:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 16 15:06:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	18473	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	30428	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	34350	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	14730	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	12617	61.10	ug/l	0.00
Spiked Amount			Recovery	=	122.20%	
35) Dibromofluoromethane	7.89	113	10995	56.32	ug/l	0.00
Spiked Amount			Recovery	=	112.64%	
50) Toluene-d8	10.33	98	43409	57.30	ug/l	0.00
Spiked Amount			Recovery	=	114.60%	
62) 4-Bromofluorobenzene	12.63	95	18484	67.66	ug/l	0.00
Spiked Amount			Recovery	=	135.32%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.13	43	2984	60.137	ug/l #	76
20) Methylene Chloride	4.93	84	419	Below Cal	#	48
25) 2-Butanone	7.16	43	123	1.842	ug/l #	51
29) Tetrahydrofuran	7.54	42	52	1.250	ug/l #	36
31) Cyclohexane	7.95	56	525	1.533	ug/l #	24
36) 1,1-Dichloropropene	7.95	75	1613	5.152	ug/l #	47
38) Carbon Tetrachloride	7.95	117	2172	7.425	ug/l #	20
41) Methacrylonitrile	7.49	41	122	1.458	ug/l #	20
46) Dibromomethane	9.47	93	225	1.956	ug/l #	37
47) Bromodichloromethane	9.68	83	289	1.019	ug/l #	24
51) 4-Methyl-2-Pentanone	10.33	43	1077	7.158	ug/l	89
55) 1,1,2-Trichloroethane	10.79	97	285	1.762	ug/l #	32
56) Ethyl methacrylate	10.66	69	122	3.442	ug/l #	1
59) 2-Hexanone	10.80	43	405	3.819	ug/l	74
65) Chlorobenzene	11.58	112	5119	7.008	ug/l #	43
71) Bromoform	12.34	173	415	2.896	ug/l #	83
75) 1,1,2,2-Tetrachloroethane	12.69	83	18869	93.739	ug/l #	27
76) 1,2,3-Trichloropropane	12.90	75	720	4.924	ug/l #	100
77) Bromobenzene	12.94	156	50188	201.510	ug/l #	1
81) trans-1,4-Dichloro-2-buten	12.48	75	235	3.555	ug/l #	1
83) tert-Butylbenzene	13.22	119	5278	7.177	ug/l #	30
86) p-Isopropyltoluene	13.51	119	3688	3.853	ug/l #	72
87) 1,3-Dichlorobenzene	13.51	146	506	1.017	ug/l #	23
90) Hexachloroethane	14.10	117	320	1.814	ug/l #	9
92) 1,2-Dibromo-3-Chloropropan	14.51	75	52	1.453	ug/l #	3


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW072618\
Data File : VW004209.D
Acq On : 25 Jul 2018 22:55
Operator : SY/AP
Sample : J4123-01
Misc : 6.94G/5ML/MSVOA_W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
RO-1-3

Quant Time: Jul 26 05:13:13 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W071618S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 16 15:06:12 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.95 min Scan# 1044

Delta R.T. -0.00 min

Lab File: VW004209.D

Acq: 25 Jul 2018 22:55

Instrument :

MSVOA_W

ClientSampleId :


RO-1-3

Tgt Ion:168 Resp: 18473

Ion Ratio Lower Upper




168 100

99 56.7 43.8 65.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VW



#16

Acetone

Concen: 60.137 ug/l

RT: 4.13 min Scan# 417

Delta R.T. 0.01 min

Lab File: VW004209.D


Acq: 25 Jul 2018 22:55

Tgt Ion: 43 Resp: 2984

Ion Ratio Lower Upper



43 100


58 17.4 24.5 36.7#



Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 58.00 (57.70 to 58.70): VW









#20
Methylene Chloride
Concen: Below Cal
RT: 4.93 min Scan# 548
Delta R.T. 0.02 min
Lab File: VW004209.D
Acq: 25 Jul 2018 22:55




Instrument :
MSVOA_W
ClientSampled :
RO-1-3


Tgt Ion: 84 Resp: 419
Ion Ratio Lower Upper
84 100
49 87.6 102.9 154.3#
51 0.0 30.9 46.3#
86 0.0 49.9 74.9#



#25
2-Butanone
Concen: 1.842 ug/l
RT: 7.16 min Scan# 914
Delta R.T. -0.02 min
Lab File: VW004209.D
Acq: 25 Jul 2018 22:55

Tgt Ion: 43 Resp: 123
Ion Ratio Lower Upper
43 100
72 0.0 19.8 29.6#





#29

Tetrahydrofuran

Concen: 1.250 ug/l

RT: 7.54 min Scan# 976

Delta R.T. 0.01 min

Lab File: VW004209.D

Acq: 25 Jul 2018 22:55

Instrument :

MSVOA_W

ClientSampleId :

RO-1-3


Tgt Ion: 42 Resp: 52

Ion Ratio Lower Upper




42 100

72 0.0 32.6 48.8#

71 0.0 30.5 45.7#



Abundance Ion 42.00 (41.70 to 42.70): VW
Ion 72.00 (71.70 to 72.70): VW
Ion 71.00 (70.70 to 71.70): VW



#31

Cyclohexane

Concen: 1.533 ug/l

RT: 7.95 min Scan# 1043

Delta R.T. -0.01 min

Lab File: VW004209.D

Acq: 25 Jul 2018 22:55


Tgt Ion: 56 Resp: 525

Ion Ratio Lower Upper



56 100


69 141.0 27.8 41.8#

84 113.0 67.2 100.8#



Abundance Ion 56.00 (55.70 to 56.70): VW
Ion 69.00 (68.70 to 69.70): VW
Ion 84.00 (83.70 to 84.70): VW









#33
 1,2-Dichloroethane-d4
 Concen: 61.102 ug/l
 RT: 8.31 min Scan# 1103
 Delta R.T. 0.01 min
 Lab File: VW004209.D
 Acq: 25 Jul 2018 22:55




Instrument :
 MSVOA_W
 ClientSampled :
 RO-1-3


Tgt Ion: 65 Resp: 12617
 Ion Ratio Lower Upper
 65 100
 67 58.0 0.0 105.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.85 min Scan# 1191
 Delta R.T. -0.00 min
 Lab File: VW004209.D
 Acq: 25 Jul 2018 22:55

Tgt Ion: 114 Resp: 30428
 Ion Ratio Lower Upper
 114 100
 63 22.6 0.0 42.8
 88 14.7 0.0 30.4





#35

Dibromofluoromethane

Concen: 56.317 ug/l

RT: 7.89 min Scan# 1034

Delta R.T. 0.01 min

Lab File: VW004209.D

Acq: 25 Jul 2018 22:55

Instrument :

MSVOA_W

ClientSampleId :

RO-1-3


Tgt Ion:113 Resp: 10995

Ion Ratio Lower Upper

113 100

111 104.1 81.3 121.9




192 19.6 15.8 23.6



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 5.152 ug/l

RT: 7.95 min Scan# 1043

Delta R.T. -0.13 min

Lab File: VW004209.D

Acq: 25 Jul 2018 22:55


Tgt Ion: 75 Resp: 1613

Ion Ratio Lower Upper

75 100

110 2.9 17.0 51.0#



77 3.2 24.9 37.3#




Abundance Ion 74.90 (74.60 to 75.60): VW

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): VW





#38

Carbon Tetrachloride

Concen: 7.425 ug/l

RT: 7.95 min Scan# 1044

Delta R.T. -0.12 min

Lab File: VW004209.D

Acq: 25 Jul 2018 22:55

Instrument :

MSVOA_W

ClientSampled :

RO-1-3


Tgt Ion: 117 Resp: 2172

Ion Ratio Lower Upper

117 100

119 9.7 75.3 112.9#




121 0.0 26.0 39.0#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#41

Methacrylonitrile

Concen: 1.458 ug/l

RT: 7.49 min Scan# 969

Delta R.T. 0.01 min

Lab File: VW004209.D

Acq: 25 Jul 2018 22:55

Tgt Ion: 41 Resp: 122


Ion Ratio Lower Upper

41 100

39 0.0 50.0 75.0#

67 0.0 60.8 91.2#

52 0.0 26.2 39.4#






Abundance Ion 41.00 (40.70 to 41.70): VW

Ion 39.00 (38.70 to 39.70): VW

Ion 67.00 (66.70 to 67.70): VW

Ion 52.00 (51.70 to 52.70): VW





#46

Dibromomethane

Concen: 1.956 ug/l

RT: 9.47 min Scan# 1293

Delta R.T. 0.01 min

Lab File: VW004209.D

Acq: 25 Jul 2018 22:55

Instrument :

MSVOA_W

ClientSampled :

RO-1-3


Tgt Ion: 93 Resp: 225

Ion Ratio Lower Upper

93 100

95 39.6 66.9 100.3#




174 24.9 80.2 120.2#



Abundance Ion 92.80 (92.50 to 93.50): VW

Ion 94.80 (94.50 to 95.50): VW

Ion 173.70 (173.40 to 174.40): V



#47

Bromodichloromethane

Concen: 1.019 ug/l

RT: 9.68 min Scan# 1327

Delta R.T. 0.03 min

Lab File: VW004209.D

Acq: 25 Jul 2018 22:55


Tgt Ion: 83 Resp: 289

Ion Ratio Lower Upper

83 100

85 0.0 52.2 78.2#



127 0.0 7.0 10.4#




Abundance Ion 82.90 (82.60 to 83.60): VW

Ion 84.90 (84.60 to 85.60): VW

Ion 126.80 (126.50 to 127.50): V





#50

Toluene-d8

Concen: 57.302 ug/l

RT: 10.33 min Scan# 1434

Delta R.T. -0.00 min

Lab File: VW004209.D

Acq: 25 Jul 2018 22:55

Instrument :

MSVOA_W

ClientSampled :


RO-1-3

Tgt Ion: 98 Resp: 43409

Ion Ratio Lower Upper




98 100

100 63.1 51.2 76.8



Abundance Ion 98.00 (97.70 to 98.70): VW

Ion 100.00 (99.70 to 100.70): VW



#51

4-Methyl-2-Pentanone

Concen: 7.158 ug/l

RT: 10.33 min Scan# 1434

Delta R.T. 0.12 min

Lab File: VW004209.D


Acq: 25 Jul 2018 22:55

Tgt Ion: 43 Resp: 1077

Ion Ratio Lower Upper



43 100


58 44.0 29.9 44.9



Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 58.00 (57.70 to 58.70): VW









#55
 1,1,2-Trichloroethane
 Concen: 1.762 ug/l
 RT: 10.79 min Scan# 1509
 Delta R.T. -0.01 min
 Lab File: VW004209.D
 Acq: 25 Jul 2018 22:55




Instrument :
 MSVOA_W
 ClientSampled :
 RO-1-3


Tgt Ion: 97 Resp: 285
 Ion Ratio Lower Upper
 97 100
 83 0.0 69.0 103.4#
 85 30.6 45.0 67.6#
 99 14.1 49.9 74.9#



#56
 Ethyl methacrylate
 Concen: 3.442 ug/l
 RT: 10.66 min Scan# 1489
 Delta R.T. 0.01 min
 Lab File: VW004209.D
 Acq: 25 Jul 2018 22:55

Tgt Ion: 69 Resp: 122
 Ion Ratio Lower Upper
 69 100
 41 0.0 56.5 84.7#
 39 675.4 30.2 45.4#









#59
2-Hexanone
Concen: 3.819 ug/l
RT: 10.80 min Scan# 1512
Delta R.T. -0.17 min
Lab File: VW004209.D
Acq: 25 Jul 2018 22:55




Instrument :
MSVOA_W
ClientSampled :
RO-1-3


Tgt Ion: 43 Resp: 405
Ion Ratio Lower Upper
43 100
58 33.8 26.0 78.0



#62
4-Bromofluorobenzene
Concen: 67.664 ug/l
RT: 12.63 min Scan# 1811
Delta R.T. 0.01 min
Lab File: VW004209.D
Acq: 25 Jul 2018 22:55

Tgt Ion: 95 Resp: 18484
Ion Ratio Lower Upper
95 100
174 65.4 0.0 160.2
176 62.6 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.63 min Scan# 1648
Delta R.T. -0.00 min
Lab File: VW004209.D
Acq: 25 Jul 2018 22:55




Instrument :
MSVOA_W
ClientSampleId :
RO-1-3


Tgt Ion:117 Resp: 34350
Ion Ratio Lower Upper
117 100
82 54.6 46.5 69.7
119 31.5 26.2 39.2



#65
Chlorobenzene
Concen: 7.008 ug/l
RT: 11.58 min Scan# 1639
Delta R.T. -0.08 min
Lab File: VW004209.D
Acq: 25 Jul 2018 22:55

Tgt Ion:112 Resp: 5119
Ion Ratio Lower Upper
112 100
114 0.0 25.2 37.8#





#71

Bromoform

Concen: 2.896 ug/l

RT: 12.34 min Scan# 1764

Delta R.T. -0.01 min

Lab File: VW004209.D

Acq: 25 Jul 2018 22:55

Instrument :

MSVOA_W

ClientSampled :

RO-1-3


Tgt Ion:173 Resp: 415

Ion Ratio Lower Upper

173 100

175 36.9 24.4 73.2

254 0.0 0.1 0.1#





Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V

12.34

Time-->



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.57 min Scan# 1965

Delta R.T. -0.00 min

Lab File: VW004209.D

Acq: 25 Jul 2018 22:55


Tgt Ion:152 Resp: 14730

Ion Ratio Lower Upper

152 100

115 63.1 29.2 87.6

150 157.1 0.0 345.2




Abundance Ion 151.90 (151.60 to 152.60): V


Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

13.57

Time-->





#75

1,1,2,2-Tetrachloroethane

Concen: 93.739 ug/l

RT: 12.69 min Scan# 1822

Delta R.T. -0.02 min

Lab File: VW004209.D

Acq: 25 Jul 2018 22:55

Instrument :

MSVOA_W

ClientSampleId :

RO-1-3


Tgt Ion: 83 Resp: 18869

Ion Ratio Lower Upper

83 100

131 2.1 5.3 15.8#




85 0.7 32.2 96.6#



Abundance Ion 82.90 (82.60 to 83.60): VW

Ion 130.80 (130.50 to 131.50): V

Ion 84.90 (84.60 to 85.60): VW



#76

1,2,3-Trichloropropane

Concen: 4.924 ug/l

RT: 12.90 min Scan# 1856

Delta R.T. 0.13 min

Lab File: VW004209.D


Acq: 25 Jul 2018 22:55

Tgt Ion: 75 Resp: 720

Ion Ratio Lower Upper



75 100


77 128.6 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VW

Ion 77.00 (76.70 to 77.70): VW





#77

Bromobenzene

Concen: 201.510 ug/l

RT: 12.94 min Scan# 1862

Delta R.T. 0.19 min

Lab File: VW004209.D

Acq: 25 Jul 2018 22:55

Instrument :

MSVOA_W

ClientSampleId :

RO-1-3


Tgt Ion:156 Resp: 50188

Ion Ratio Lower Upper

156 100

77 1.8 96.2 288.6#




158 1.1 48.5 145.5#



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VW

Ion 157.80 (157.50 to 158.50): V



#81

trans-1,4-Dichloro-2-butene

Concen: 3.555 ug/l

RT: 12.48 min Scan# 1787

Delta R.T. -0.04 min

Lab File: VW004209.D

Acq: 25 Jul 2018 22:55


Tgt Ion: 75 Resp: 235

Ion Ratio Lower Upper

75 100

53 246.8 77.6 116.4#



89 57.4 36.6 54.8#




Abundance Ion 74.90 (74.60 to 75.60): VW

Ion 53.00 (52.70 to 53.70): VW

Ion 88.90 (88.60 to 89.60): VW









#83
tert-Butylbenzene
Concen: 7.177 ug/l
RT: 13.22 min Scan# 1908
Delta R.T. 0.01 min
Lab File: VW004209.D
Acq: 25 Jul 2018 22:55




Instrument :
MSVOA_W
ClientSampled :
RO-1-3


Tgt Ion:119 Resp: 5278
Ion Ratio Lower Upper
119 100
91 2.3 33.2 99.6#
134 3.2 13.0 38.9#



#86
p-Isopropyltoluene
Concen: 3.853 ug/l
RT: 13.51 min Scan# 1955
Delta R.T. -0.00 min
Lab File: VW004209.D
Acq: 25 Jul 2018 22:55

Tgt Ion:119 Resp: 3688
Ion Ratio Lower Upper
119 100
134 0.0 13.0 38.9#
91 24.5 11.8 35.3









#87
 1,3-Dichlorobenzene
 Concen: 1.017 ug/l
 RT: 13.51 min Scan# 1956
 Delta R.T. 0.01 min
 Lab File: VW004209.D
 Acq: 25 Jul 2018 22:55




Instrument :
 MSVOA_W
 ClientSampled :
 RO-1-3


Tgt Ion:146 Resp: 506
 Ion Ratio Lower Upper
 146 100
 111 83.2 19.9 59.6#
 148 0.0 31.9 95.5#



#90
 Hexachloroethane
 Concen: 1.814 ug/l
 RT: 14.10 min Scan# 2053
 Delta R.T. -0.00 min
 Lab File: VW004209.D
 Acq: 25 Jul 2018 22:55

Tgt Ion:117 Resp: 320
 Ion Ratio Lower Upper
 117 100
 201 0.0 40.3 120.9#





#92
 1,2-Dibromo-3-Chloropropane
 Concen: 1.453 ug/l
 RT: 14.51 min Scan# 2119
 Delta R.T. 0.01 min
 Lab File: VW004209.D
 Acq: 25 Jul 2018 22:55

Instrument :
 MSVOA_W
 ClientSampleId :
 RO-1-3

Tgt Ion: 75 Resp: 52
 Ion Ratio Lower Upper
 75 100
 155 0.0 39.4 118.1#
 157 0.0 51.8 155.4#

