

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072621\
 Data File : VW019609.D
 Acq On : 26 Jul 2021 20:43
 Operator : SY/VA
 Sample : M3164-01
 Misc : 5.48g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JCEN0

Quant Time: Jul 27 01:13:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jul 27 01:09:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.847	114	468390	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.633	117	454799	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	201683	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	89399	13.471	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	53.880%
7) Chloroethane-d5	2.897	69	86769	18.335	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.320%
11) 1,1-Dichloroethene-d2	4.025	63	165970	12.961	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	51.840%
21) 2-Butanone-d5	7.079	46	72311	37.541	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.080%
24) Chloroform-d	7.652	84	284460	21.979	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.920%
26) 1,2-Dichloroethane-d4	8.311	65	155945	21.957	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.840%
32) Benzene-d6	8.274	84	558325	20.623	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	82.480%
36) 1,2-Dichloropropane-d6	9.274	67	177442	21.029	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	84.120%
41) Toluene-d8	10.323	98	487763	19.983	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	79.920%
43) trans-1,3-Dichloroprop...	10.579	79	66299	18.347	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	73.400%
47) 2-Hexanone-d5	10.926	63	55379	35.376	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	70.760%
56) 1,1,2,2-Tetrachloroeth...	12.694	84	138109	20.256	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	81.040%
66) 1,2-Dichlorobenzene-d4	13.853	152	156775	22.032	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	88.120%
Target Compounds						Qvalue
16) Methylene chloride	4.927	84	17763	1.902	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072621\
Data File : VW019609.D
Acq On : 26 Jul 2021 20:43
Operator : SY/VA
Sample : M3164-01
Misc : 5.48g/10.0mL/MSVOA_W/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JCEN0

Quant Time: Jul 27 01:13:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070921SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jul 27 01:09:44 2021
Response via : Initial Calibration

