

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072621\
 Data File : VW019597.D
 Acq On : 26 Jul 2021 15:32
 Operator : SY/VA
 Sample : M3162-13
 Misc : 3.91g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGEX0

Quant Time: Jul 27 01:11:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jul 27 01:09:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	409308	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	365821	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	143939	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	80535	13.887	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	55.560%		
7) Chloroethane-d5	2.898	69	77643	18.774	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	75.080%		
11) 1,1-Dichloroethene-d2	4.026	63	143807	12.852	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	51.400%		
21) 2-Butanone-d5	7.086	46	70460	41.860	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	83.720%		
24) Chloroform-d	7.653	84	240935	21.304	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	85.200%		
26) 1,2-Dichloroethane-d4	8.311	65	135552	21.841	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	87.360%		
32) Benzene-d6	8.281	84	467067	21.449	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	85.800%		
36) 1,2-Dichloropropane-d6	9.275	67	152053	22.403	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	89.600%		
41) Toluene-d8	10.323	98	396827	20.212	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	80.840%		
43) trans-1,3-Dichloroprop...	10.579	79	53980	18.571	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	74.280%		
47) 2-Hexanone-d5	10.927	63	54387	43.192	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	86.380%		
56) 1,1,2,2-Tetrachloroeth...	12.695	84	114420	20.863	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	83.440%		
66) 1,2-Dichlorobenzene-d4	13.853	152	105251	20.725	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	82.880%		
Target Compounds					Qvalue	
13) Acetone	4.135	43	49714	40.083	ug/L	87
22) 2-Butanone	7.184	43	10855	5.656	ug/L	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072621\
Data File : VW019597.D
Acq On : 26 Jul 2021 15:32
Operator : SY/VA
Sample : M3162-13
Misc : 3.91g/10.0mL/MSVOA_W/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGEX0

Quant Time: Jul 27 01:11:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070921SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jul 27 01:09:44 2021
Response via : Initial Calibration









#13
Acetone
Concen: 40.083 ug/L
RT: 4.135 min Scan# 366
Delta R.T. -0.012 min
Lab File: VW019597.D
Acq: 26 Jul 2021 15:32

Instrument :
MSVOA_W
ClientSampleId :
BGEX0

Tgt Ion: 43 Resp: 49714
Ion Ratio Lower Upper
43 100
58 33.6 0.0 53.8



#22
2-Butanone
Concen: 5.656 ug/L
RT: 7.184 min Scan# 866
Delta R.T. 0.006 min
Lab File: VW019597.D
Acq: 26 Jul 2021 15:32

Tgt Ion: 43 Resp: 10855
Ion Ratio Lower Upper
43 100
72 25.5 14.9 44.7

