

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072721\
 Data File : VW019627.D
 Acq On : 27 Jul 2021 17:11
 Operator : SY/VA
 Sample : M3163-02
 Misc : 4.53g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGEX2

Quant Time: Jul 28 00:51:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jul 28 00:48:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.847	114	473420	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.633	117	450414	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	206837	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	135405	20.186	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	80.760%	
7) Chloroethane-d5	2.897	69	100049	20.916	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	83.680%	
11) 1,1-Dichloroethene-d2	4.025	63	194227	15.007	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	60.040%	
21) 2-Butanone-d5	7.079	46	68094	34.976	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	69.960%	
24) Chloroform-d	7.652	84	270927	20.711	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	82.840%	
26) 1,2-Dichloroethane-d4	8.311	65	148812	20.730	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	82.920%	
32) Benzene-d6	8.280	84	559667	20.874	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	83.480%	
36) 1,2-Dichloropropane-d6	9.274	67	170816	20.441	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	81.760%	
41) Toluene-d8	10.329	98	508315	21.028	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	84.120%	
43) trans-1,3-Dichloroprop...	10.579	79	66652	18.624	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	74.480%	
47) 2-Hexanone-d5	10.926	63	53245	34.344	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	68.680%	
56) 1,1,2,2-Tetrachloroeth...	12.694	84	128980	19.101	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	76.400%	
66) 1,2-Dichlorobenzene-d4	13.852	152	161208	22.090	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	88.360%	
Target Compounds						
					Qvalue	
13) Acetone	4.135	43	19010	13.252	ug/L	89
14) Carbon disulfide	4.391	76	8120	0.410	ug/L	100
42) Toluene	10.390	91	11889	0.405	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW072721\
Data File : VW019627.D
Acq On : 27 Jul 2021 17:11
Operator : SY/VA
Sample : M3163-02
Misc : 4.53g/10.0mL/MSVOA_W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGEX2

Quant Time: Jul 28 00:51:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM070921SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jul 28 00:48:17 2021
Response via : Initial Calibration

