

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072820\
 Data File : VW015979.D
 Acq On : 28 Jul 2020 13:18
 Operator : SY/VA
 Sample : L3458-11
 Misc : 4.60G/10ML/MSVOA W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GAY37

Quant Time: Jul 29 05:10:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM072020S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 29 04:38:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	527325	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	496329	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	219985	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	136499	18.34	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.36%
7) Chloroethane-d5	2.90	69	111051	18.38	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.52%
10) 1,1-Dichloroethene-d2	4.02	63	214460	14.26	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.04%
20) 2-Butanone-d5	7.09	46	83649	42.80	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.60%
24) Chloroform-d	7.66	84	287447	19.41	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	77.64%
26) 1,2-Dichloroethane-d4	8.31	65	157635	19.39	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	77.56%
29) Benzene-d6	8.28	84	580774	19.22	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	76.88%
33) 1,2-Dichloropropane-d6	9.28	67	179417	19.81	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	79.24%
37) Toluene-d8	10.33	98	509164	18.61	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	74.44%
38) trans-1,3-Dichloropropene-	10.58	79	73254	18.98	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	75.92%
39) 2-Hexanone-d5	10.93	63	66736	44.25	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.50%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	146791	21.09	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	84.36%
61) 1,2-Dichlorobenzene-d4	13.86	152	181467	21.16	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	84.64%

Target Compounds

					Ovalue
13) Acetone	4.14	43	17325	10.555	ug/L 98
15) Methyl Acetate	4.68	43	2503	0.809	ug/L # 77
22) cis-1,2-Dichloroethene	7.17	96	487300	65.067	ug/L 96
31) 1,1,1-Trichloroethane	7.88	97	9610	0.829	ug/L 99
35) Trichloroethene	9.09	95	2585593	324.240	ug/L 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW072820\
Data File : VW015979.D
Acq On : 28 Jul 2020 13:18
Operator : SY/VA
Sample : L3458-11
Misc : 4.60G/10ML/MSVOA W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GAY37

Quant Time: Jul 29 05:10:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM072020S.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 29 04:38:00 2020
Response via : Initial Calibration

