

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072918\
 Data File : VW004273.D
 Acq On : 27 Jul 2018 15:31
 Operator : SY/AP
 Sample : J4169-07
 Misc : 5.63G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BEY62

Quant Time: Jul 27 23:17:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM072618S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 27 23:17:19 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	396770	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	342430	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	121442	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.34	65	100534	25.17	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	100.68%
7) Chloroethane-d5	2.89	69	71249	24.52	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	98.08%
10) 1,1-Dichloroethene-d2	4.00	63	105567	11.07	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	44.28%#
20) 2-Butanone-d5	7.08	46	110176	48.72	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.44%
24) Chloroform-d	7.65	84	248292	24.44	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.76%
26) 1,2-Dichloroethane-d4	8.31	65	155202	25.40	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	101.60%
29) Benzene-d6	8.28	84	494037	27.22	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	108.88%
33) 1,2-Dichloropropane-d6	9.27	67	167655	27.77	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	111.08%
37) Toluene-d8	10.33	98	439236	25.61	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	102.44%
38) trans-1,3-Dichloropropene-	10.58	79	63935	23.34	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	93.36%
39) 2-Hexanone-d5	10.93	63	81813	51.58	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	103.16%
48) 1,1,2,2-Tetrachloroethane-	12.70	84	158145	25.66	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	102.64%
61) 1,2-Dichlorobenzene-d4	13.86	152	128630	28.33	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	113.32%

Target Compounds						Ovalue
13) Acetone	4.12	43	4860	2.13	ug/L	96
22) cis-1,2-Dichloroethene	7.17	96	24701	4.60	ug/L	97
35) Trichloroethene	9.09	95	24289	4.69	ug/L	99
47) Tetrachloroethene	10.87	164	60556	14.62	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW072918\
Data File : VW004273.D
Acq On : 27 Jul 2018 15:31
Operator : SY/AP
Sample : J4169-07
Misc : 5.63G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BEY62

Quant Time: Jul 27 23:17:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM072618S.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 27 23:17:19 2018
Response via : Initial Calibration

