

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072918\
 Data File : VW004277.D
 Acq On : 27 Jul 2018 17:31
 Operator : SY/AP
 Sample : J4205-02
 Misc : 4.23G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JJ1K9

Quant Time: Jul 28 00:04:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM072618S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 27 23:17:19 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	407389	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.64	117	381244	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	162443	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.34	65	130334	31.78	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	127.12%
7) Chloroethane-d5	2.89	69	90393	30.30	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	121.20%
10) 1,1-Dichloroethene-d2	4.00	63	215480	22.01	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	88.04%
20) 2-Butanone-d5	7.08	46	99899	43.03	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.06%
24) Chloroform-d	7.65	84	277372	26.59	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	106.36%
26) 1,2-Dichloroethane-d4	8.31	65	169939	27.09	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	108.36%
29) Benzene-d6	8.27	84	563503	27.89	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	111.56%
33) 1,2-Dichloropropane-d6	9.28	67	185356	27.58	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	110.32%
37) Toluene-d8	10.33	98	521470	27.30	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	109.20%
38) trans-1,3-Dichloropropene-	10.58	79	78959	25.89	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	103.56%
39) 2-Hexanone-d5	10.93	63	77534	43.90	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.80%
48) 1,1,2,2-Tetrachloroethane-	12.70	84	162535	23.68	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	94.72%
61) 1,2-Dichlorobenzene-d4	13.87	152	171079	28.17	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	112.68%

Target Compounds					Ovalue
13) Acetone	4.12	43	4745	2.02 ug/L	98
16) Methylene chloride	4.91	84	13816	2.42 ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072918\
Data File : VW004277.D
Acq On : 27 Jul 2018 17:31
Operator : SY/AP
Sample : J4205-02
Misc : 4.23G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JJ1K9

Quant Time: Jul 28 00:04:12 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM072618S.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 27 23:17:19 2018
Response via : Initial Calibration

